



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100504

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	01/28/2019	10:30
2	CAL2	1	Water	NA	NA	22.1	7.01	01/28/2019	10:31
3	CAL3	1	Water	NA	NA	22.1	10.04	01/28/2019	10:32
4	ICV	1	Water	NA	NA	21.2	7.00	01/28/2019	10:33
5	CCV1	1	Water	NA	NA	21.2	2.01	01/28/2019	10:34
6	K1263-01	1	Water	NA	NA	20.0	4.74	01/28/2019	10:35
7	K1263-01DUP	1	Water	NA	NA	20.1	4.75	01/28/2019	10:36
8	CCV2	1	Water	NA	NA	22.1	12.01	01/28/2019	10:37

2B 100504

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1263

WorkList ID : 121674

Date : 1/28/2019 10:20:04 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2019	Water	K1263-01 -A	pH	Cool 4 deg C	STER14	L31	MINI-CELL	01/24/2019	SM 4500-H B

Date/Time 01-28-19 10:25 AM
Received by: JB
Relinquished by: CB

Date/Time 01-28-19 11:00 AM
Received by: CB
Relinquished by: JB