

LB100505


INSTRUMENT ID : CV1

7470A

Linear

 μ Abs.:

43119



A= 0.0000e+000

B= 2.3079e-004

C= -5.9118e-003

Rho= 0.9999101

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	30	0.000	30				
0.2	0.200	0.200	-0.000	891	0.0 %	891				
2.5	2.500	2.436	-0.064	10580	0.0 %	10580				
5.0	5.000	5.078	0.078	22026	0.0 %	22026				
7.5	7.500	7.540	0.040	32696	0.0 %	32696				
10.0	10.000	9.946	-0.054	43119	0.0 %	43119	0			

LB100505 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Jan 2019 12:20:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	30	-	0.0000	7470A	PPB	28 Jan 2019 12:21:30	S	Std
0.2 - 1	50-2	891	-	0.2000	7470A	PPB	28 Jan 2019 12:23:38	S	Std
2.5 - 1	52-5	10580	-	2.5000	7470A	PPB	28 Jan 2019 12:25:47	S	Std
5.0 - 1	55-0	22026	-	5.0000	7470A	PPB	28 Jan 2019 12:27:56	S	Std
7.5 - 1	57-5	32696	-	7.5000	7470A	PPB	28 Jan 2019 12:30:04	S	Std
10.0 - 1	510-0	43119	-	10.0000	7470A	PPB	28 Jan 2019 12:32:11	S	Std
ICV82 - 1	ICV82	17713	4.0821	-	7470A	PPB	28 Jan 2019 12:35:14	U	SMPL
ICB82 - 1	ICB82	-26	-0.0119	-	7470A	PPB	28 Jan 2019 12:37:21	U	SMPL
CCV24 - 1	CCV24	22315	5.1443	-	7470A	PPB	28 Jan 2019 12:39:28	U	SMPL
CCB24 - 1	CCB24	-49	-0.0172	-	7470A	PPB	28 Jan 2019 12:41:34	U	SMPL
CRA - 1	CRA	874	0.1958	-	7470A	PPB	28 Jan 2019 12:43:41	U	SMPL
HighStd - 1	HighStd	44418	10.2455	-	7470A	PPB	28 Jan 2019 12:45:48	U	SMPL
ChkStd - 1	ChkStd	31900	7.3564	-	7470A	PPB	28 Jan 2019 12:47:54	U	SMPL
PB116635BL - 1	PBW	-73	-0.0228	-	7470A	PPB	28 Jan 2019 12:50:03	U	SMPL
PB116635BS - 1	LCSW	17629	4.0628	-	7470A	PPB	28 Jan 2019 12:56:01	U	SMPL
K1015-14 - 1	SU-04-012419-A	-28	-0.0124	-	7470A	PPB	28 Jan 2019 12:58:09	U	SMPL
K1015-14DUP - 1	SU-04-012419-ADUP	9	-0.0038	-	7470A	PPB	28 Jan 2019 13:00:16	U	SMPL
K1015-14MS - 1	SU-04-012419-AMS	16674	3.8423	-	7470A	PPB	28 Jan 2019 13:02:25	U	SMPL
K1015-14MSD - 1	SU-04-012419-AMSD	16795	3.8703	-	7470A	PPB	28 Jan 2019 13:04:33	U	SMPL
K1015-16 - 1	SU-04-012419-B	-28	-0.0124	-	7470A	PPB	28 Jan 2019 13:06:40	U	SMPL
CCV25 - 1	CCV25	22117	5.0986	-	7470A	PPB	28 Jan 2019 13:08:46	U	SMPL
CCB25 - 1	CCB25	-12	-0.0087	-	7470A	PPB	28 Jan 2019 13:10:53	U	SMPL
K1228-02 - 1	COMP-1	-14	-0.0091	-	7470A	PPB	28 Jan 2019 13:13:00	U	SMPL
K1231-02 - 1	RBR-200029	16	-0.0022	-	7470A	PPB	28 Jan 2019 13:15:07	U	SMPL
K1233-03 - 1	SANDHILL-RO-1-8	16	-0.0022	-	7470A	PPB	28 Jan 2019 13:17:14	U	SMPL
K1234-04 - 1	EWP	18	-0.0018	-	7470A	PPB	28 Jan 2019 13:19:21	U	SMPL
K1238-10 - 1	S1	10	-0.0036	-	7470A	PPB	28 Jan 2019 13:21:29	U	SMPL
K1238-20 - 1	S2	15	-0.0024	-	7470A	PPB	28 Jan 2019 13:23:37	U	SMPL
K1238-30 - 1	S3	19	-0.0015	-	7470A	PPB	28 Jan 2019 13:25:45	U	SMPL
K1238-40 - 1	S4	20	-0.0013	-	7470A	PPB	28 Jan 2019 13:27:53	U	SMPL
K1238-50 - 1	S5	30	0.0010	-	7470A	PPB	28 Jan 2019 13:30:02	U	SMPL
K1242-01 - 1	DRUM-SET-1	-11	-0.0085	-	7470A	PPB	28 Jan 2019 13:32:09	U	SMPL
CCV26 - 1	CCV26	22414	5.1671	-	7470A	PPB	28 Jan 2019 13:34:16	U	SMPL
CCB26 - 1	CCB26	-19	-0.0103	-	7470A	PPB	28 Jan 2019 13:36:23	U	SMPL
K1242-02 - 1	DRUM-SET-2	7	-0.0043	-	7470A	PPB	28 Jan 2019 13:38:32	U	SMPL
K1242-03 - 1	DRUM-SET-3	-25	-0.0117	-	7470A	PPB	28 Jan 2019 13:40:39	U	SMPL
K1242-04 - 1	DRUM-SET-4	21	-0.0011	-	7470A	PPB	28 Jan 2019 13:42:46	U	SMPL
K1242-05 - 1	DRUM-SET-5	23	-0.0006	-	7470A	PPB	28 Jan 2019 13:44:54	U	SMPL
K1242-06 - 1	DRUM-OILY-STONE	8	-0.0041	-	7470A	PPB	28 Jan 2019 13:47:04	U	SMPL
PB116635TB - 1	PB116608TB	35	0.0022	-	7470A	PPB	28 Jan 2019 13:49:16	U	SMPL
K1015-14LX5 - 1		29	0.0008	-	7470A	PPB	28 Jan 2019 13:51:38	U	SMPL
K1015-14A - 1		18842	4.3427	-	7470A	PPB	28 Jan 2019 13:53:47	U	SMPL
CCV27 - 1	CCV27	22709	5.2352	-	7470A	PPB	28 Jan 2019 13:55:56	U	SMPL
CCB27 - 1	CCB27	-14	-0.0091	-	7470A	PPB	28 Jan 2019 13:58:04	U	SMPL