



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100517

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	01/28/2019	09:25
2	K1250-05		1	6	>100.0	>212.0	>212.0	01/28/2019	09:50
3	K1256-01		1	8	>100.0	>212.0	>212.0	01/28/2019	10:20
4	K1256-01DUP		1	8	>100.0	>212.0	>212.0	01/28/2019	10:50
5	K1256-02		1	9	>100.0	>212.0	>212.0	01/28/2019	11:25
6	K1256-03		1	8	>100.0	>212.0	>212.0	01/28/2019	11:45
7	K1256-04		1	9	>100.0	>212.0	>212.0	01/28/2019	12:15
8	K1256-05		1	8	>100.0	>212.0	>212.0	01/28/2019	12:40
9	K1256-06		1	7	>100.0	>212.0	>212.0	01/28/2019	13:02
10	K1260-01		1	7	>100.0	>212.0	>212.0	01/28/2019	13:40
11	K1260-02		1	6	>100.0	>212.0	>212.0	01/28/2019	14:05
12	K1260-03		1	7	>100.0	>212.0	>212.0	01/28/2019	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB100577

WorkList Name : FP12819

WorkList ID : 121670

Date : 1/28/2019 8:49:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2019	Water	K1250-05	Flash Point	Cool 4 deg C	PSEG01	K31	DRUM-1-3	01/25/2019	1010A
01/30/2019	Water	K1256-01	Flash Point	Cool 4 deg C	PSEG01	K31	30093	01/25/2019	1010A
01/30/2019	Water	K1256-02	Flash Point	Cool 4 deg C	PSEG01	K31	30175	01/25/2019	1010A
01/30/2019	Water	K1256-03	Flash Point	Cool 4 deg C	PSEG01	K31	30176	01/25/2019	1010A
01/30/2019	Water	K1256-04	Flash Point	Cool 4 deg C	PSEG01	K31	50138	01/25/2019	1010A
01/30/2019	Water	K1256-05	Flash Point	Cool 4 deg C	PSEG01	K31	50139	01/25/2019	1010A
01/30/2019	Water	K1256-06	Flash Point	Cool 4 deg C	PSEG01	K31	50140	01/25/2019	1010A
02/08/2019	Water	K1260-01	Flash Point	Cool 4 deg C	TULL02	EL-EAST-OF-JDHS		01/25/2019	1010A
02/08/2019	Water	K1260-02	Flash Point	Cool 4 deg C	TULL02	MHC05		01/25/2019	1010A
02/08/2019	Water	K1260-03	Flash Point	Cool 4 deg C	TULL02	MH-J09		01/25/2019	1010A

Date/Time 01/28/19 8:50 AM
 Received by: KS
 Relinquished by: CP

Date/Time 01/28/19 14:45
 Received by: CP
 Relinquished by: KS