



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100519

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	01/28/2019	17:35
2	CAL2	1	Water	NA	NA	22.2	7.01	01/28/2019	17:36
3	CAL3	1	Water	NA	NA	22.1	10.02	01/28/2019	17:37
4	ICV	1	Water	NA	NA	21.1	6.99	01/28/2019	17:38
5	CCV1	1	Water	NA	NA	22.2	2.01	01/28/2019	17:39
6	K1008-04	1	Solid	20.02	20	20.5	8.20	01/28/2019	17:40
7	K1008-04DUP	1	Solid	20.03	20	20.6	8.25	01/28/2019	17:41
8	K1010-08	1	Solid	20.05	20	19.4	7.73	01/28/2019	17:42
9	K1010-10	1	Solid	20.02	20	19.9	7.12	01/28/2019	17:43
10	K1016-12	1	Solid	20.01	20	20.1	9.81	01/28/2019	17:44
11	K1265-01	1	Solid	20.00	20	18.0	6.65	01/28/2019	17:45
12	CCV2	1	Water	NA	NA	21.1	12.01	01/28/2019	17:46

23 100519

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrsivity k1264

WorkList ID : 121685

Date : 1/28/2019 1:19:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2019	Solid	K1008-04	Corrosivity	Cool 4 deg C	PSEG05		HD-01-012819	01/28/2019	9045D
01/28/2019	Solid	K1010-08	Corrosivity	Cool 4 deg C	PSEG05		OK-03-012819-A	01/28/2019	9045D
01/28/2019	Solid	K1010-10	Corrosivity	Cool 4 deg C	PSEG05		OK-03-012819-B	01/28/2019	9045D
01/28/2019	Solid	K1016-12	Corrosivity	Cool 4 deg C	PSEG05		TR-05-012819	01/28/2019	9045D
01/25/2019	Solid	K1265-01	Corrosivity	Cool 4 deg C	CLEA11	L41	OWS-3(OUTSIDE)	01/25/2019	9045D

Date/Time 01-28-19 5:10 PM
 Received by: SP
 Relinquished by: SP

Date/Time 01-28-19 6:30 PM
 Received by: SP
 Relinquished by: SP