


LB100523
INSTRUMENT ID: CV1

7470A

Linear



A= 0.0000e+000

B= 2.2436e-004

C= -7.0833e-003

Rho= 0.9998123

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	19	0.000	19				
0.2	0.200	0.185	-0.015	854	0.0 %	854				
2.5	2.500	2.466	-0.034	11025	0.0 %	11025				
5.0	5.000	5.027	0.027	22436	0.0 %	22436				
7.5	7.500	7.631	0.131	34046	0.0 %	34046				
10.0	10.000	9.894	-0.106	44129	0.0 %	44129	0			

LB100523 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Jan 2019 09:15:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	19	-	0.00007470A	PPB		29 Jan 2019 09:18:48	S	Std
0.2 - 1	50.2	854	-	0.20007470A	PPB		29 Jan 2019 09:20:57	S	Std
2.5 - 1	52.5	11025	-	2.50007470A	PPB		29 Jan 2019 09:23:06	S	Std
5.0 - 1	55.0	22436	-	5.00007470A	PPB		29 Jan 2019 09:25:15	S	Std
7.5 - 1	57.5	34046	-	7.50007470A	PPB		29 Jan 2019 09:27:23	S	Std
10.0 - 1	510.0	44129	-	10.00007470A	PPB		29 Jan 2019 09:29:30	S	Std
ICV83 - 1	ICV83	16726	3.7456	-7470A	PPB		29 Jan 2019 09:33:04	U	SMPL
ICB83 - 1	ICB83	-48	-0.0179	-7470A	PPB		29 Jan 2019 09:35:11	U	SMPL
CCV28 - 1	CCV28	21576	4.8337	-7470A	PPB		29 Jan 2019 09:37:18	U	SMPL
CCB28 - 1	CCB28	-55	-0.0194	-7470A	PPB		29 Jan 2019 09:39:25	U	SMPL
CRA - 1	CRA	1185	0.2588	-7470A	PPB		29 Jan 2019 09:41:32	U	SMPL
HighStd - 1	HighStd	43667	9.7900	-7470A	PPB		29 Jan 2019 09:43:39	U	SMPL
ChkStd - 1	ChkStd	31568	7.0755	-7470A	PPB		29 Jan 2019 09:45:46	U	SMPL
PB116666BL - 1	PBW	-61	-0.0208	-7470A	PPB		29 Jan 2019 09:47:54	U	SMPL
PB116666BS - 1	LCSW	17046	3.8174	-7470A	PPB		29 Jan 2019 09:50:01	U	SMPL
K1011-10 - 1	OR-01-012519-A	-44	-0.0170	-7470A	PPB		29 Jan 2019 09:52:09	U	SMPL
K1011-10DUP - 1	OR-01-012519-ADUP	11	-0.0046	-7470A	PPB		29 Jan 2019 09:54:18	U	SMPL
K1011-10MS - 1	OR-01-012519-AMS	16387	3.6695	-7470A	PPB		29 Jan 2019 09:56:27	U	SMPL
K1011-10MSD - 1	OR-01-012519-AMSD	16427	3.6785	-7470A	PPB		29 Jan 2019 09:58:35	U	SMPL
K1011-12 - 1	OR-01-012519-B	-49	-0.0181	-7470A	PPB		29 Jan 2019 10:00:42	U	SMPL
CCV29 - 1	CCV29	22013	4.9317	-7470A	PPB		29 Jan 2019 10:02:49	U	SMPL
CCB29 - 1	CCB29	-64	-0.0214	-7470A	PPB		29 Jan 2019 10:04:56	U	SMPL
K1250-03 - 1	DRUM-6	37	0.0012	-7470A	PPB		29 Jan 2019 10:07:03	U	SMPL
K1254-02 - 1	SOIL-1	19	-0.0028	-7470A	PPB		29 Jan 2019 10:09:09	U	SMPL
K1254-04 - 1	SOIL-2	20	-0.0026	-7470A	PPB		29 Jan 2019 10:11:17	U	SMPL
K1255-02 - 1	RBR200036	6	-0.0057	-7470A	PPB		29 Jan 2019 10:13:24	U	SMPL
K1261-02 - 1	EL-EAST-OF-JDHS	21	-0.0024	-7470A	PPB		29 Jan 2019 10:15:32	U	SMPL
K1261-04 - 1	MH-C05	31	-0.0001	-7470A	PPB		29 Jan 2019 10:17:40	U	SMPL
PB116667BL - 1	PBW	13	-0.0042	-7470A	PPB		29 Jan 2019 10:19:48	U	SMPL
PB116667BS - 1	LCSW	15939	3.5690	-7470A	PPB		29 Jan 2019 10:21:57	U	SMPL
K1250-05 - 1	DRUM-1-3	-14	-0.0102	-7470A	PPB		29 Jan 2019 10:24:06	U	SMPL
K1256-01 - 1	30093	18	-0.0030	-7470A	PPB		29 Jan 2019 10:26:13	U	SMPL
CCV30 - 1	CCV30	22054	4.9409	-7470A	PPB		29 Jan 2019 10:28:20	U	SMPL
CCB30 - 1	CCB30	-38	-0.0156	-7470A	PPB		29 Jan 2019 10:30:27	U	SMPL
K1256-01DUP - 1	30093DUP	22	-0.0021	-7470A	PPB		29 Jan 2019 10:32:34	U	SMPL
K1256-01MS - 1	30093MS	16585	3.7139	-7470A	PPB		29 Jan 2019 10:34:42	U	SMPL
K1256-01MSD - 1	30093MSD	16229	3.6341	-7470A	PPB		29 Jan 2019 10:36:48	U	SMPL
K1256-02 - 1	30175	252	0.0495	-7470A	PPB		29 Jan 2019 10:38:56	U	SMPL
K1256-03 - 1	30176	-38	-0.0156	-7470A	PPB		29 Jan 2019 10:41:03	U	SMPL
K1256-04 - 1	50138	108	0.0171	-7470A	PPB		29 Jan 2019 10:43:11	U	SMPL
K1256-05 - 1	50139	16	-0.0035	-7470A	PPB		29 Jan 2019 10:45:19	U	SMPL
K1256-06 - 1	50140	8	-0.0053	-7470A	PPB		29 Jan 2019 10:47:28	U	SMPL
PB116666TB - 1	PB116639TB	41	0.0021	-7470A	PPB		29 Jan 2019 10:49:36	U	SMPL
PB116667TB - 1	PB116641TB	31	-0.0001	-7470A	PPB		29 Jan 2019 10:51:44	U	SMPL
CCV31 - 1	CCV31	22068	4.9441	-7470A	PPB		29 Jan 2019 10:53:51	U	SMPL
CCB31 - 1	CCB31	-41	-0.0163	-7470A	PPB		29 Jan 2019 10:55:59	U	SMPL
K1011-10LX5 - 1		10	-0.0048	-7470A	PPB		29 Jan 2019 10:58:06	U	SMPL
K1011-10A - 1		17143	3.8391	-7470A	PPB		29 Jan 2019 11:00:14	U	SMPL
K1256-01LX5 - 1		-27	-0.0131	-7470A	PPB		29 Jan 2019 11:02:21	U	SMPL
K1256-01A - 1		17783	3.9827	-7470A	PPB		29 Jan 2019 11:04:29	U	SMPL
CCV32 - 1	CCV32	22231	4.9807	-7470A	PPB		29 Jan 2019 11:06:36	U	SMPL
CCB32 - 1	CCB32	-57	-0.0199	-7470A	PPB		29 Jan 2019 11:08:44	U	SMPL