



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100525

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/29/2019	10:30
2	CAL2	1	Water	NA	NA	21.1	7.00	01/29/2019	10:31
3	CAL3	1	Water	NA	NA	21.1	10.03	01/29/2019	10:32
4	ICV	1	Water	NA	NA	21.2	7.01	01/29/2019	10:33
5	CCV1	1	Water	NA	NA	21.2	2.02	01/29/2019	10:36
6	K1274-01	1	Water	NA	NA	16.2	4.92	01/29/2019	10:37
7	K1274-01DUP	1	Water	NA	NA	16.3	4.93	01/29/2019	10:38
8	CCV2	1	Water	NA	NA	21.2	12.01	01/29/2019	10:39

LB100525

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1274 WorkList ID : 121709 Date : 1/29/2019 10:16:31 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2019	Water	K1274-01-1A	pH	Cool 4 deg C	STER11	L51	POOLWATER	01/15/2019	9940C

SM4500-HB
JP

Date/Time 01-29-18 10:20 AM
Received by: JO
Relinquished by: CQ

Date/Time 01-29-18 11:10 AM
Received by: CQ
Relinquished by: JP