



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100529

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/29/2019	13:15
2	CAL2	1	Water	NA	NA	20.1	7.01	01/29/2019	13:16
3	CAL3	1	Water	NA	NA	20.1	10.02	01/29/2019	13:17
4	ICV	1	Water	NA	NA	20.1	7.01	01/29/2019	13:18
5	CCV1	1	Water	NA	NA	20.3	2.02	01/29/2019	13:20
6	K1272-01	1	Solid	20.02	20	20.8	7.40	01/29/2019	13:21
7	K1272-01DUP	1	Solid	20.03	20	20.7	7.41	01/29/2019	13:22
8	CCV2	1	Water	NA	NA	20.1	12.02	01/29/2019	13:25

LB 100529

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1272

WorkList ID : 121717

Date : 1/29/2019 1:04:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2019	Solid	K1272-01-C	pH	Cool 4 deg C	SEVE02	L41	TS-01-012819	01/28/2019	9045D

Date/Time 01-29-18 11:10 PM
Received by: JS
Relinquished by: CS

Date/Time 01-29-18 2:35 PM
Received by: CS
Relinquished by: AB