

LB100538


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58019



A= 0.0000e+000

B= 1.7185e-004

C= 2.8477e-002

Rho= 0.9999035

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	-32	0.000	-32				
0.2	0.200	0.221	0.021	1118	0.0 %	1118				
2.5	2.500	2.403	-0.097	13815	0.0 %	13815				
5.0	5.000	5.071	0.071	29343	0.0 %	29343				
7.5	7.500	7.484	-0.016	43381	0.0 %	43381				
10.0	10.000	9.999	-0.001	58019	0.0 %	58019				

LB100538 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Jan 2019 15:09:53

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	100	-32	-	0.00007471B	PPB	29 Jan 2019 15:11:20	S	Std
0.2 - 1	50.2	1118	-	0.20007471B	PPB	29 Jan 2019 15:13:38	S	Std
2.5 - 1	52.5	13815	-	2.50007471B	PPB	29 Jan 2019 15:15:56	S	Std
5.0 - 1	55.0	29343	-	5.00007471B	PPB	29 Jan 2019 15:18:14	S	Std
7.5 - 1	57.5	43381	-	7.50007471B	PPB	29 Jan 2019 15:20:31	S	Std
10.0 - 1	510.0	58019	-	10.00007471B	PPB	29 Jan 2019 15:31:27	S	Std
ICV84 - 1	ICV84	22000	3.8092	-7471B	PPB	29 Jan 2019 15:34:50	U	SMPL
ICB84 - 1	ICB84	-78	0.0151	-7471B	PPB	29 Jan 2019 15:37:07	U	SMPL
CCV33 - 1	CCV33	28843	4.9852	-7471B	PPB	29 Jan 2019 15:39:23	U	SMPL
CCB33 - 1	CCB33	-95	0.0122	-7471B	PPB	29 Jan 2019 15:41:39	U	SMPL
CRA - 1	CRA	1107	0.2187	-7471B	PPB	29 Jan 2019 15:43:56	U	SMPL
HighStd - 1	HighStd	58507	10.0830	-7471B	PPB	29 Jan 2019 15:46:12	U	SMPL
ChkStd - 1	ChkStd	41270	7.1208	-7471B	PPB	29 Jan 2019 15:48:28	U	SMPL
PB116675BL - 1	PBS	-110	0.0096	-7471B	PPB	29 Jan 2019 15:50:46	U	SMPL
PB116675BS - 1	LCSS	22358	3.8707	-7471B	PPB	29 Jan 2019 16:01:31	U	SMPL
K1010-07 - 1	OK-03-012819-A	1036	0.2065	-7471B	PPB	29 Jan 2019 16:03:48	U	SMPL
K1010-09 - 1	OK-03-012819-B	1697	0.3201	-7471B	PPB	29 Jan 2019 16:06:05	U	SMPL
K1011-09 - 1	OR-01-012519-A	2820	0.5131	-7471B	PPB	29 Jan 2019 16:08:23	U	SMPL
K1011-11 - 1	OR-01-012519-B	3210	0.5801	-7471B	PPB	29 Jan 2019 16:10:41	U	SMPL
K1016-11 - 1	TR-05-012819	938	0.1897	-7471B	PPB	29 Jan 2019 16:12:57	U	SMPL
CCV34 - 1	CCV34	29252	5.0555	-7471B	PPB	29 Jan 2019 16:54:11	U	SMPL
CCB34 - 1	CCB34	-58	0.0185	-7471B	PPB	29 Jan 2019 16:56:27	U	SMPL
K1250-02 - 1	DRUM-6	3099	0.5610	-7471B	PPB	29 Jan 2019 16:58:44	U	SMPL
K1254-01 - 1	SOIL-1	1447	0.2771	-7471B	PPB	29 Jan 2019 17:01:00	U	SMPL
K1254-03 - 1	SOIL-2	1736	0.3268	-7471B	PPB	29 Jan 2019 17:03:16	U	SMPL
K1255-01 - 1	RBR200036	769	0.1606	-7471B	PPB	29 Jan 2019 17:05:32	U	SMPL
K1258-01 - 1	FK-BPM-01-AA01-BB01-AA01A-AA01B	1598	0.3031	-7471B	PPB	29 Jan 2019 17:07:51	U	SMPL
K1258-01DUP - 1	FK-BPM-01-AA01-BB01-AA01A-AA01BDUP	1616	0.3062	-7471B	PPB	29 Jan 2019 17:10:08	U	SMPL
K1258-01MS - 1	FK-BPM-01-AA01-BB01-AA01A-AA01BMS	21896	3.7913	-7471B	PPB	29 Jan 2019 17:12:25	U	SMPL
K1258-01MSD - 1	FK-BPM-01-AA01-BB01-AA01A-AA01BMSD	22043	3.8166	-7471B	PPB	29 Jan 2019 17:14:43	U	SMPL
K1259-01 - 1	FK-BPM-01-A01-B01-A01A-A01B	1564	0.2973	-7471B	PPB	29 Jan 2019 17:17:01	U	SMPL
K1261-01 - 1	EL-EAST-OF-JDHS	8519	1.4925	-7471B	PPB	29 Jan 2019 17:19:18	U	SMPL
CCV35 - 1	CCV35	28600	4.9434	-7471B	PPB	29 Jan 2019 17:21:35	U	SMPL
CCB35 - 1	CCB35	-64	0.0175	-7471B	PPB	29 Jan 2019 17:23:51	U	SMPL
K1261-03 - 1	MH-C05	5070	0.8998	-7471B	PPB	29 Jan 2019 17:26:08	U	SMPL
K1258-01LX5 - 1		298	0.0797	-7471B	PPB	29 Jan 2019 17:28:24	U	SMPL
K1258-01A - 1		24452	4.2306	-7471B	PPB	29 Jan 2019 17:30:41	U	SMPL
CCV36 - 1	CCV36	29047	5.0202	-7471B	PPB	29 Jan 2019 17:32:57	U	SMPL
CCB36 - 1	CCB36	-72	0.0161	-7471B	PPB	29 Jan 2019 17:35:14	U	SMPL