



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100542

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/29/2019	17:15
2	CAL2	1	Water	NA	NA	21.2	7.00	01/29/2019	17:16
3	CAL3	1	Water	NA	NA	21.2	10.02	01/29/2019	17:17
4	ICV	1	Water	NA	NA	21.1	6.99	01/29/2019	17:18
5	CCV1	1	Water	NA	NA	21.2	2.01	01/29/2019	17:20
6	K1277-03	1	Water	NA	NA	10.2	7.98	01/29/2019	17:21
7	K1277-03DUP	1	Water	NA	NA	10.3	7.99	01/29/2019	17:22
8	K1282-35	1	Water	NA	NA	11.2	6.28	01/29/2019	17:24
9	CCV2	1	Water	NA	NA	21.1	12.02	01/29/2019	17:25

13 100542

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1282

WorkList ID : 121724

Date : 1/29/2019 4:14:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2019	Water	K1277-03 ~ 1A	pH	Cool 4 deg C	PSEG01	K31	FRAC-TANK-1-2	01/29/2019	9040C
01/29/2019	Water	K1282-35 ~ 1A	pH	Cool 4 deg C	PSEG01	K31	TPW1	01/29/2019	9040C

Date/Time 01-29-19 4:13:52 PM
Received by: 30
Relinquished by: 30

Date/Time 01-29-19 6:20 PM
Received by: 30
Relinquished by: 30