

LB100544


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

53140



A= 0.0000e+000

B= 1.8896e-004

C= -3.4429e-002

Rho= 0.9997293

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	154	0.000	154				
0.2	0.200	0.173	-0.027	1096	0.0 %	1096				
2.5	2.500	2.476	-0.024	13283	0.0 %	13283				
5.0	5.000	5.169	0.169	27539	0.0 %	27539				
7.5	7.500	7.381	-0.119	39244	0.0 %	39244				
10.0	10.000	10.007	0.007	53140	0.0 %	53140				

LB100544 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 30 Jan 2019 08:58:01					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	154	-	0.0000	7471B	PPB	30 Jan 2019 11:25:49	S	Std
0.2 - 1	50-2	1096	-	0.2000	7471B	PPB	30 Jan 2019 11:28:06	S	Std
2.5 - 1	52-5	13283	-	2.5000	7471B	PPB	30 Jan 2019 11:30:24	S	Std
5.0 - 1	55-0	27539	-	5.0000	7471B	PPB	30 Jan 2019 11:32:41	S	Std
7.5 - 1	57-5	39244	-	7.5000	7471B	PPB	30 Jan 2019 11:34:57	S	Std
10.0 - 1	510-0	53140	-	10.0000	7471B	PPB	30 Jan 2019 11:37:15	S	Std
ICV85 - 1	ICV85	21097	3.9520	-	7471B	PPB	30 Jan 2019 12:03:45	U	SMPL
ICB85 - 1	ICB85	340	0.0298	-	7471B	PPB	30 Jan 2019 12:05:59	U	SMPL
CCV37 - 1	CCV37	27949	5.2468	-	7471B	PPB	30 Jan 2019 12:08:16	U	SMPL
CCB37 - 1	CCB37	331	0.0281	-	7471B	PPB	30 Jan 2019 12:10:30	U	SMPL
CRA - 1	CRA	1052	0.1644	-	7471B	PPB	30 Jan 2019 12:12:47	U	SMPL
HighStd - 1	HighStd	53861	10.1431	-	7471B	PPB	30 Jan 2019 12:17:16	U	SMPL
ChkStd - 1	ChkStd	36546	6.8712	-	7471B	PPB	30 Jan 2019 12:19:31	U	SMPL
PB116716BL - 1	PBS	-46	-0.0431	-	7471B	PPB	30 Jan 2019 12:21:49	U	SMPL
PB116716BS - 1	LCSS	21607	4.0484	-	7471B	PPB	30 Jan 2019 12:29:09	U	SMPL
K1008-03 - 1	HD-01-012819	520	0.0638	-	7471B	PPB	30 Jan 2019 12:31:25	U	SMPL
K1008-03DUP - 1	HD-01-012819DUP	578	0.0748	-	7471B	PPB	30 Jan 2019 12:33:42	U	SMPL
K1008-03MS - 1	HD-01-012819MS	18857	3.5288	-	7471B	PPB	30 Jan 2019 12:36:00	U	SMPL
K1008-03MSD - 1	HD-01-012819MSD	18877	3.5325	-	7471B	PPB	30 Jan 2019 12:38:17	U	SMPL
K1271-01 - 1	OILY-DEBRIS	438	0.0483	-	7471B	PPB	30 Jan 2019 12:40:33	U	SMPL
CCV38 - 1	CCV38	27412	5.1453	-	7471B	PPB	30 Jan 2019 12:42:50	U	SMPL
CCB38 - 1	CCB38	-7	-0.0358	-	7471B	PPB	30 Jan 2019 12:45:05	U	SMPL
K1272-01 - 1	TS-01-012819	5179	0.9442	-	7471B	PPB	30 Jan 2019 12:47:22	U	SMPL
K1281-01 - 1	SB-16(35.5-36)	398	0.0408	-	7471B	PPB	30 Jan 2019 12:49:37	U	SMPL
K1281-02 - 1	SB-15(33.5-34)	414	0.0438	-	7471B	PPB	30 Jan 2019 12:51:52	U	SMPL
K1282-01 - 1	TP-1	13041	2.4298	-	7471B	PPB	30 Jan 2019 12:54:08	U	SMPL
K1282-05 - 1	TP-2	36245	6.8144	-	7471B	PPB	30 Jan 2019 12:56:24	U	SMPL
K1282-09 - 1	TP-3	18131	3.3916	-	7471B	PPB	30 Jan 2019 12:58:41	U	SMPL
K1282-13 - 1	TP-4	17442	3.2614	-	7471B	PPB	30 Jan 2019 13:00:59	U	SMPL
K1282-17 - 1	TP-5	32199	6.0498	-	7471B	PPB	30 Jan 2019 13:03:18	U	SMPL
K1282-21 - 1	TP-6	34608	6.5050	-	7471B	PPB	30 Jan 2019 13:05:36	U	SMPL
K1282-25 - 1	TP-7	40280	7.5768	-	7471B	PPB	30 Jan 2019 13:07:53	U	SMPL
CCV39 - 1	CCV39	27820	5.2224	-	7471B	PPB	30 Jan 2019 13:10:11	U	SMPL
CCB39 - 1	CCB39	-27	-0.0395	-	7471B	PPB	30 Jan 2019 13:12:29	U	SMPL
K1282-29 - 1	TP-8-AB	1837	0.3127	-	7471B	PPB	30 Jan 2019 13:14:46	U	SMPL
K1008-03LX5 - 1		169	-0.0025	-	7471B	PPB	30 Jan 2019 13:17:02	U	SMPL
K1008-03A - 1		21964	4.1159	-	7471B	PPB	30 Jan 2019 13:19:17	U	SMPL
CCV40 - 1	CCV40	27254	5.1154	-	7471B	PPB	30 Jan 2019 13:21:33	U	SMPL
CCB40 - 1	CCB40	-9	-0.0361	-	7471B	PPB	30 Jan 2019 13:23:50	U	SMPL