



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100551

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	01/30/2019	09:45
2	K1277-04		1	9	>100.0	>212.0	>212.0	01/30/2019	10:20
3	K1277-04DUP		1	9	>100.0	>212.0	>212.0	01/30/2019	10:50
4	K1282-36		1	8	>100.0	>212.0	>212.0	01/30/2019	11:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB100557

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP13019 WorkList ID : 121749 Date : 1/30/2019 8:59:51 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/01/2019	Water	K1277-04 C	Flash Point	Cool 4 deg C	PSEG01	K31	FRAC-TANK-1-2	01/29/2019	1010A
02/04/2019	Water	K1282-36 C	Flash Point	Cool 4 deg C	PSEG01	K31	TPW1	01/29/2019	1010A

Date/Time 01/30/19 9:15 AM
Received by: KS
Relinquished by: CP

Date/Time 01/30/19 11:30 AM
Received by: CP
Relinquished by: KS