

LB100552


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37814



A= 0.0000e+000

B= 2.6620e-004

C= -1.3099e-002

Rho= 0.9999547

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	118	0.000	118				
0.2	0.200	0.201	0.001	804	0.0 %	804				
2.5	2.500	2.514	0.014	9492	0.0 %	9492				
5.0	5.000	4.971	-0.029	18723	0.0 %	18723				
7.5	7.500	7.443	-0.057	28010	0.0 %	28010				
10.0	10.000	10.053	0.053	37814	0.0 %	37814				

LB100552 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jan 2019 13:48:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	118	-	0.00007470A	PPB		30 Jan 2019 13:49:24	S	Std
0.2 - 1	50-2	804	-	0.20007470A	PPB		30 Jan 2019 13:51:32	S	Std
2.5 - 1	52-5	9492	-	2.50007470A	PPB		30 Jan 2019 13:53:40	S	Std
5.0 - 1	55-0	18723	-	5.00007470A	PPB		30 Jan 2019 13:55:48	S	Std
7.5 - 1	57-5	28010	-	7.50007470A	PPB		30 Jan 2019 13:57:55	S	Std
10.0 - 1	510-D	37814	-	10.00007470A	PPB		30 Jan 2019 14:14:49	S	Std
ICV86 - 1	ICV86	13671	3.6261	-7470A	PPB		30 Jan 2019 14:17:39	U	SMPL
ICB86 - 1	ICB86	66	0.0045	-7470A	PPB		30 Jan 2019 14:19:46	U	SMPL
CCV41 - 1	CCV41	18515	4.9156	-7470A	PPB		30 Jan 2019 14:21:52	U	SMPL
CCB41 - 1	CCB41	58	0.0023	-7470A	PPB		30 Jan 2019 14:23:57	U	SMPL
CRA - 1	CRA	779	0.1943	-7470A	PPB		30 Jan 2019 14:26:03	U	SMPL
HighStd - 1	HighStd	37577	9.9899	-7470A	PPB		30 Jan 2019 14:28:09	U	SMPL
ChkStd - 1	ChkStd	26720	7.0998	-7470A	PPB		30 Jan 2019 14:30:15	U	SMPL
PB116703BL - 1	PBW	48	-0.0003	-7470A	PPB		30 Jan 2019 14:32:22	U	SMPL
PB116703BS - 1	LCSW	14334	3.8026	-7470A	PPB		30 Jan 2019 14:34:30	U	SMPL
K1008-04 - 1	HD-01-012819	241	0.0511	-7470A	PPB		30 Jan 2019 14:36:36	U	SMPL
K1010-08 - 1	OK-03-012819-A	165	0.0308	-7470A	PPB		30 Jan 2019 14:38:44	U	SMPL
K1010-08DUP - 1	OK-03-012819-ADUP	177	0.0340	-7470A	PPB		30 Jan 2019 14:40:51	U	SMPL
K1010-08MS - 1	OK-03-012819-AMS	14119	3.7454	-7470A	PPB		30 Jan 2019 14:43:00	U	SMPL
K1010-08MSD - 1	OK-03-012819-AMSD	14510	3.8495	-7470A	PPB		30 Jan 2019 14:45:06	U	SMPL
CCV42 - 1	CCV42	18594	4.9366	-7470A	PPB		30 Jan 2019 14:47:12	U	SMPL
CCB42 - 1	CCB42	57	0.0021	-7470A	PPB		30 Jan 2019 14:49:18	U	SMPL
K1010-10 - 1	OK-03-012819-B	97	0.0127	-7470A	PPB		30 Jan 2019 14:51:24	U	SMPL
K1016-12 - 1	TR-05-012819	237	0.0500	-7470A	PPB		30 Jan 2019 14:53:30	U	SMPL
K1264-01 - 1	RH-CONCRETE	372	0.0859	-7470A	PPB		30 Jan 2019 14:55:36	U	SMPL
K1265-01 - 1	OWS-3(OUTSIDE)	207	0.0420	-7470A	PPB		30 Jan 2019 14:57:43	U	SMPL
PB116715BL - 1	PBW	138	0.0236	-7470A	PPB		30 Jan 2019 15:03:02	U	SMPL
PB116715BS - 1	LCSW	14315	3.7976	-7470A	PPB		30 Jan 2019 15:05:09	U	SMPL
K1250-04 - 1	DRUM-1-3	3728	0.9793	-7470A	PPB		30 Jan 2019 15:07:16	U	SMPL
K1277-03 - 1	FRAC-TANK-1-2	807	0.2017	-7470A	PPB		30 Jan 2019 15:09:24	U	SMPL
K1277-03DUP - 1	FRAC-TANK-1-2DUP	903	0.2273	-7470A	PPB		30 Jan 2019 15:11:33	U	SMPL
K1277-03MS - 1	FRAC-TANK-1-2MS	15976	4.2397	-7470A	PPB		30 Jan 2019 15:13:40	U	SMPL
CCV43 - 1	CCV43	18369	4.8767	-7470A	PPB		30 Jan 2019 15:15:46	U	SMPL
CCB43 - 1	CCB43	83	0.0090	-7470A	PPB		30 Jan 2019 15:17:52	U	SMPL
K1277-03MSD - 1	FRAC-TANK-1-2MSD	16026	4.2530	-7470A	PPB		30 Jan 2019 15:19:59	U	SMPL
K1282-35 - 1	TPW1	12364	3.2782	-7470A	PPB		30 Jan 2019 15:22:06	U	SMPL
PB116703TB - 1	PB116676TB	72	0.0061	-7470A	PPB		30 Jan 2019 15:24:12	U	SMPL
K1010-08LX5 - 1		138	0.0236	-7470A	PPB		30 Jan 2019 15:26:19	U	SMPL
K1010-08A - 1		15371	4.0787	-7470A	PPB		30 Jan 2019 15:28:25	U	SMPL
K1277-03LX5 - 1		202	0.0407	-7470A	PPB		30 Jan 2019 15:30:32	U	SMPL
K1277-03A - 1		16420	4.3579	-7470A	PPB		30 Jan 2019 15:32:39	U	SMPL
CCV44 - 1	CCV44	18369	4.8767	-7470A	PPB		30 Jan 2019 15:34:47	U	SMPL
CCB44 - 1	CCB44	68	0.0050	-7470A	PPB		30 Jan 2019 15:36:55	U	SMPL