

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB100559

ANALYST: ketankumar
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP70959
IODINE SOLUTION .025N 1L	W2410
Starch Solution, 4L	W2277

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB116731BL		1	50	50	2.00	0.00	1.80	1.80	0.20	0.00	0.00	01/30/2019	14:45
2	K1277-04		1	50	50	2.00	1.80	3.48	1.68	0.32	0.12	0.96	01/30/2019	14:48
3	K1277-04DUP		1	50	50	2.00	3.48	5.18	1.70	0.30	0.10	0.80	01/30/2019	14:52
4	K1282-36		1	50	50	2.00	0.00	1.60	1.60	0.40	0.20	1.60	01/30/2019	14:55

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume