


LB100562

245.1

INSTRUMENT ID : CV1

 μ Abs.:

36821



Linear

A= 0.0000e+000

B= 2.7125e-004

C= -5.1206e-002

Rho= 0.9999394

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	32	0.000	32				
0.05	0.050	0.053	0.003	383	0.0 %	383				
0.2	0.200	0.193	-0.007	901	0.0 %	901				
2.5	2.500	2.536	0.036	9540	0.0 %	9540				
5.0	5.000	5.008	0.008	18650	0.0 %	18650				
7.5	7.500	7.566	0.066	28083	0.0 %	28083				
10.0	10.000	9.936	-0.064	36821	0.0 %	36821				

LB100562 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 30 Jan 2019 16:26:59					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	32	-	0.0000	245.1	PPB	30 Jan 2019 16:49:47	S	Std
0.05 - 1	50.05	383	-	0.0500	245.1	PPB	30 Jan 2019 16:52:05	S	Std
0.2 - 1	50.2	901	-	0.2000	245.1	PPB	30 Jan 2019 16:54:22	S	Std
2.5 - 1	52.5	9540	-	2.5000	245.1	PPB	30 Jan 2019 16:56:39	S	Std
5.0 - 1	55.0	18650	-	5.0000	245.1	PPB	30 Jan 2019 16:58:55	S	Std
7.5 - 1	57.5	28083	-	7.5000	245.1	PPB	30 Jan 2019 17:01:11	S	Std
10.0 - 1	510.0	36821	-	10.0000	245.1	PPB	30 Jan 2019 17:03:27	S	Std
ICV87 - 1	ICV87	14824	3.9698	-245.1		PPB	30 Jan 2019 17:06:29	U	SMPL
ICB87 - 1	ICB87	-10	-0.0539	-245.1		PPB	30 Jan 2019 17:11:23	U	SMPL
CCV45 - 1	CCV45	18469	4.9585	-245.1		PPB	30 Jan 2019 17:13:38	U	SMPL
CCB45 - 1	CCB45	-61	-0.0678	-245.1		PPB	30 Jan 2019 17:15:53	U	SMPL
CRA - 1	CRA	359	0.0462	-245.1		PPB	30 Jan 2019 17:18:08	U	SMPL
HighStd - 1	HighStd	36612	9.8797	-245.1		PPB	30 Jan 2019 17:20:23	U	SMPL
ChkStd - 1	ChkStd	25706	6.9215	-245.1		PPB	30 Jan 2019 17:22:38	U	SMPL
PB116699BL - 1	PBW	-14	-0.0550	-245.1		PPB	30 Jan 2019 17:24:55	U	SMPL
PB116699BS - 1	LCSW	14735	3.9456	-245.1		PPB	30 Jan 2019 17:27:12	U	SMPL
K1229-01 - 1	012319-TIPC-F-D-B	50	-0.0376	-245.1		PPB	30 Jan 2019 17:29:29	U	SMPL
K1229-01DUP - 1	012319-TIPC-F-D-BDUP	84	-0.0284	-245.1		PPB	30 Jan 2019 17:31:46	U	SMPL
K1229-01MS - 1	012319-TIPC-F-D-BMS	16678	4.4727	-245.1		PPB	30 Jan 2019 17:34:03	U	SMPL
K1229-01MSD - 1	012319-TIPC-F-D-BMSD	16454	4.4119	-245.1		PPB	30 Jan 2019 17:36:20	U	SMPL
K1229-02 - 1	012319-TIPC-F-D-S	78	-0.0300	-245.1		PPB	30 Jan 2019 17:46:37	U	SMPL
CCV46 - 1	CCV46	18381	4.9346	-245.1		PPB	30 Jan 2019 17:48:53	U	SMPL
CCB46 - 1	CCB46	-102	-0.0789	-245.1		PPB	30 Jan 2019 17:51:08	U	SMPL
K1229-03 - 1	012319-TIPC-F-U-B	151	-0.0102	-245.1		PPB	30 Jan 2019 17:53:24	U	SMPL
K1229-04 - 1	012319-TIPC-F-U-S	113	-0.0206	-245.1		PPB	30 Jan 2019 17:55:39	U	SMPL
K1260-01 - 1	EL-EAST-OF-JDHS	2382	0.5949	-245.1		PPB	30 Jan 2019 17:57:55	U	SMPL
K1260-02 - 1	MH-C05	151	-0.0102	-245.1		PPB	30 Jan 2019 18:00:10	U	SMPL
K1260-03 - 1	MH-J09	192	0.0009	-245.1		PPB	30 Jan 2019 18:02:27	U	SMPL
K1266-01 - 1	012519-SIDI-E-D-B	150	-0.0105	-245.1		PPB	30 Jan 2019 18:04:43	U	SMPL
K1266-02 - 1	012519-SIDI-E-D-M	122	-0.0181	-245.1		PPB	30 Jan 2019 18:06:59	U	SMPL
K1266-03 - 1	012519-SIDI-E-D-S	94	-0.0257	-245.1		PPB	30 Jan 2019 18:09:16	U	SMPL
K1266-04 - 1	012519-SIDI-E-U-B	83	-0.0287	-245.1		PPB	30 Jan 2019 18:11:33	U	SMPL
K1266-05 - 1	012519-SIDI-E-U-M	99	-0.0244	-245.1		PPB	30 Jan 2019 18:13:49	U	SMPL
CCV47 - 1	CCV47	18328	4.9202	-245.1		PPB	30 Jan 2019 18:16:05	U	SMPL
CCB47 - 1	CCB47	-86	-0.0745	-245.1		PPB	30 Jan 2019 18:18:20	U	SMPL
K1266-06 - 1	012519-SIDI-E-U-S	93	-0.0260	-245.1		PPB	30 Jan 2019 18:20:37	U	SMPL
K1266-07 - 1	012519-TIPC-F-D-B	95	-0.0254	-245.1		PPB	30 Jan 2019 18:22:52	U	SMPL
K1266-08 - 1	012519-TIPC-F-D-S	108	-0.0219	-245.1		PPB	30 Jan 2019 18:25:08	U	SMPL
K1266-09 - 1	012519-TIPC-F-U-B	89	-0.0271	-245.1		PPB	30 Jan 2019 18:27:24	U	SMPL
K1266-10 - 1	012519-TIPC-F-U-S	87	-0.0276	-245.1		PPB	30 Jan 2019 18:29:39	U	SMPL
K1266-11 - 1	012519-DBRI-F-D-M	99	-0.0244	-245.1		PPB	30 Jan 2019 18:31:56	U	SMPL
K1266-12 - 1	012519-DBRI-F-U-M	92	-0.0263	-245.1		PPB	30 Jan 2019 18:34:12	U	SMPL
K1229-01LX5 - 1		8	-0.0490	-245.1		PPB	30 Jan 2019 18:36:30	U	SMPL
K1229-01A - 1		15874	4.2546	-245.1		PPB	30 Jan 2019 18:38:46	U	SMPL
CCV48 - 1	CCV48	18537	4.9769	-245.1		PPB	30 Jan 2019 18:41:02	U	SMPL
CCB48 - 1	CCB48	-59	-0.0672	-245.1		PPB	30 Jan 2019 18:43:18	U	SMPL