



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB100564

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	01/30/2019	15:05
2	K1276-01		1	8	>100.0	>212.0	>212.0	01/30/2019	15:40
3	K1276-01DUP		1	8	>100.0	>212.0	>212.0	01/30/2019	16:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 100564

WorkList Name : FP13019-1

WorkList ID : 121763

Date : 1/30/2019 2:31:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2019	Solid	K1276-01	Flash Point	Cool 4 deg C	PSEG01	K31	WASTE-OIL	01/29/2019	1010A

Date/Time 01/30/19 2:35 PM

Received by: KS

Relinquished by: CP

Date/Time 01/30/19 16:15

Received by: CP

Relinquished by: KS