



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB100568
Analysis Date: 01/31/2019
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 01/31/2019
Extraction IN Time: 11:20
Extraction OUT Time: 11:50

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB100568BL	LB100568BL	WATER	2	1000	100	2.1066	2.1066	3.02	2.1066	2.1066	0.0000	0
2	LB100568BS	LB100568BS	WATER	2	1000	100	2.8743	2.8743	3.01	2.8931	2.8931	0.0188	18.8
3	K1294-01	STORM-DRAIN-1	WATER	2	1000	100	2.8793	2.8793	3.03	2.8796	2.8796	0.0003	0.3
4	K1294-01DUP	STORM-DRAIN-1DUP	WATER	2	1000	100	3.7461	3.7461	3.02	3.7465	3.7465	0.0004	0.4
5	K1294-02	STORM-DRAIN-2	WATER	2	1000	100	3.0692	3.0692	3.03	3.0703	3.0703	0.0011	1.1
6	K1294-02MS	STORM-DRAIN-2	WATER	2	1000	100	3.0718	3.0718	3.04	3.0928	3.0928	0.0210	21



QC Batch# LB100568 Test: TPH Analysis Date: 01/31/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2478
pH Paper 0-14	W2435
Sodium Sulfate	EP1922
1:1 HCL	WP69312
Silica Gel	W2466
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP69314
LCSWD	NA	NA
MS/MSD	10.00 ML	WP69313

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>13:21</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>12:20</u>		
Bal Check Time:	<u>11:35</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>13:40</u>
		Out Time:	<u>13:20</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>14:36</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>14:15</u>		
Bal Check Time:	<u>15:35</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>15:10</u>
		Out Time:	<u>14:35</u>		

LB 100568

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K1294 WorkList ID : 121782 Date : 1/31/2019 11:00:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2019	Water	K1294-01 E, F	TPH	Conc H2SO4	KWAR01	L61	STORM-DRAIN-1	01/30/2019	1664A
02/06/2019	Water	K1294-02	TPH	Conc H2SO4	KWAR01	L61	STORM-DRAIN-2	01/30/2019	1664A

E, F

Date/Time01-31-1911:10 AM

Received by: J

Relinquished by: J

Date/Time01-31-1911:00 AM

Received by: C

Relinquished by: J