



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100571

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/31/2019	13:15
2	CAL2	1	Water	NA	NA	20.1	7.01	01/31/2019	13:16
3	CAL3	1	Water	NA	NA	20.1	10.02	01/31/2019	13:17
4	ICV	1	Water	NA	NA	20.2	6.99	01/31/2019	13:18
5	CCV1	1	Water	NA	NA	20.1	2.02	01/31/2019	13:19
6	K1297-01	1	Solid	20.02	20	18.4	7.16	01/31/2019	13:20
7	K1297-01DUP	1	Solid	20.03	20	18.5	7.17	01/31/2019	13:21
8	K1297-04	1	Solid	20.04	20	18.8	7.46	01/31/2019	13:22
9	K1297-05	1	Solid	20.05	20	18.7	7.71	01/31/2019	13:24
10	K1297-06	1	Solid	20.03	20	18.7	7.64	01/31/2019	13:25
11	CCV2	1	Water	NA	NA	20.1	12.02	01/31/2019	13:26

LB 10057A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1297

WorkList ID : 121787

Date : 1/31/2019 1:01:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/30/2019	Solid	K1297-01-10	pH	Cool 4 deg C	WEST10	L62	A0C17-SSI	01/30/2019	9045D
01/30/2019	Solid	K1297-04-10	pH	Cool 4 deg C	WEST10	L62	DUP-01-30-19	01/30/2019	9045D
01/30/2019	Solid	K1297-05-10	pH	Cool 4 deg C	WEST10	L62	A0C17-SS2	01/30/2019	9045D
01/30/2019	Solid	K1297-06-10	pH	Cool 4 deg C	WEST10	L62	A0C4-SS1	01/30/2019	9045D

Date/Time 01/31/19 11:10 AM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 01/31/19 2:00 PM
Received by: [Signature]
Relinquished by: [Signature]