



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 2/13/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 01/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:05
Out Date: 02/01/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: %Solids-Oven

QC:LB100579

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1011-13	OR-02-013119-A	1	1.14	9.82	8.89	89.3
K1011-15	OR-02-013119-B	2	1.16	9.7	8.88	90.4
K1297-01	A0C17-SS1	3	1.18	9.94	7.41	71.1
K1297-02	K1297-01MS	4	1.18	9.94	7.41	71.1
K1297-03	K1297-01MSD	5	1.18	9.94	7.41	71.1
K1297-04	DUP-01-30-19	6	1.16	9.81	6.72	64.3
K1297-05	A0C17-SS2	7	1.17	9.76	8.36	83.7
K1297-06	A0C4-SS1	8	1.17	9.69	9.32	95.7
K1299-01	0035-COMP-1	9	1.18	9.91	5.43	48.7
K1299-02	0036-COMP-2	10	1.16	9.82	5.24	47.1
K1299-03	0037-COMP-3	11	1.18	9.97	5.43	48.4
K1299-04	0041-FIELD-DUP	12	1.1	9.63	4.83	43.7
K1299-05	0038-AMND-COMP-1	13	1.18	9.95	5.85	53.2
K1299-06	0039-AMND-COMP-2	14	1.19	9.83	5.46	49.4
K1299-07	0040-AMND-COMP-3	15	1.17	9.61	5.6	52.5
K1299-08	SPLP-LEACH-AMND-COMP-1	16	1.00	2.00	1.00	0.0
K1299-09	SPLP-LEACH-AMND-COMP-2	17	1.00	2.00	1.00	0.0
K1299-10	SPLP-LEACH-AMND-COMP-3	18	1.00	2.00	1.00	0.0
K1301-01	OILY-DEBRIS-DRUM	19	1.00	2.00	2.00	100.0
K1303-01	131-A-E	20	1.00	2.00	2.00	100.0
K1303-02	131-F-J	21	1.00	2.00	2.00	100.0
K1303-03	131-K-O	22	1.00	2.00	2.00	100.0
K1303-04	131-P-T	23	1.00	2.00	2.00	100.0
K1303-05	131-U-X	24	1.00	2.00	2.00	100.0
K1303-06	131-Y-Z	25	1.00	2.00	2.00	100.0
K1308-01	SB-17 (34.2-34.8)	26	1.18	9.97	9.1	90.1
K1308-02	SB-18 (34.5-35)	27	1.14	9.87	8.8	87.7

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$