



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100580

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/31/2019	17:15
2	CAL2	1	Water	NA	NA	20.2	7.00	01/31/2019	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	01/31/2019	17:17
4	ICV	1	Water	NA	NA	20.1	7.00	01/31/2019	17:18
5	CCV1	1	Water	NA	NA	20.2	2.01	01/31/2019	17:20
6	K1011-14	1	Solid	20.03	20	17.9	7.93	01/31/2019	17:21
7	K1011-14DUP	1	Solid	20.04	20	18.6	8.11	01/31/2019	17:22
8	K1011-16	1	Solid	20.05	20	18.6	8.11	01/31/2019	17:23
9	CCV2	1	Water	NA	NA	20.1	12.02	01/31/2019	17:25

18100580

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k1011 WorkList ID : 121794 Date : 1/31/2019 4:30:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2019	Solid	K1011-14	Corrosivity	Cool 4 deg C	PSEG05		OR-02-013119-A	01/31/2019	9045D
01/31/2019	Solid	K1011-16	Corrosivity	Cool 4 deg C	PSEG05		OR-02-013119-B	01/31/2019	9045D

Date/Time 01-31-19 4:35 PM
Received by: Jo
Relinquished by: [Signature]

Date/Time 01-31-19 6:00 PM
Received by: [Signature]
Relinquished by: JP