



Extraction and Analytical Summary Report

Analysis Method: 1664A

Test: Oil and Grease

Run Number: LB100587

Analysis Date: 02/01/2019

BalanceID: WC SC-3

OvenID: EX ONEN-2

ANALYST: jignesh

REVIEWED BY: apatel

Extraction Date: 02/01/2019

Extraction IN Time: 07:40

Extraction OUT Time: 08:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB100587BL	LB100587BL	WATER	2	1000	100	2.5630	2.5630	0	2.5630	2.5630	0.0000	0
2	LB100587BS	LB100587BS	WATER	2	1000	100	2.1403	2.1403	0	2.1590	2.1590	0.0187	18.7
3	LB100587BSD	LB100587BSD	WATER	2	1000	100	2.8414	2.8414	0	2.8603	2.8603	0.0189	18.9
4	K1293-01	001M-WILLETS-PT-BLVD (J	WATER	2	1000	100	3.0694	3.0694	0	3.0792	3.0792	0.0098	9.8
5	K1293-02	002M-35TH-AVE (JAN)	WATER	2	1000	100	3.0303	3.0303	0	3.0406	3.0406	0.0103	10.3



QC Batch# LB100587 Test: Oil and Grea Analysis Date: 02/01/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2478
pH Paper 0-14	W2435
Sodium Sulfate	EP1922
1:1 HCL	WP69312
Silica Gel	NA
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP69314
LCSWD	5.00 ML	WP69313
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:36</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>08:30</u>		
Bal Check Time:	<u>08:00</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:00</u>
		Out Time:	<u>09:35</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:01</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>10:30</u>		
Bal Check Time:	<u>12:05</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:30</u>
		Out Time:	<u>11:00</u>		

FB10058F

WORKLIST(Hardcopy Internal Chain)

WorkList Name : o & g k1293

WorkList ID : 121796

Date : 2/1/2019 7:29:24 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/07/2019	Water	K1293-01 ✓	Oil and Grease	Conc H2SO4 tc	TULL01	L43	001M-WILLETS-PT-BLVD(JAN)	01/30/2019	1664A
02/07/2019	Water	K1293-02 ✓	Oil and Grease	Conc H2SO4 tc	TULL01	L43	002M-35TH-AVE(JAN)	01/30/2019	1664A

Date/Time 02-01-19 7:35 AM
 Received by: Jo
 Relinquished by: CB

Date/Time 02-01-19 10:00 AM
 Received by: Jo
 Relinquished by: Jo