



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100592

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/01/2019	13:25
2	CAL2	1	Water	NA	NA	20.1	6.99	02/01/2019	13:26
3	CAL3	1	Water	NA	NA	20.1	10.02	02/01/2019	13:27
4	ICV	1	Water	NA	NA	20.1	7.00	02/01/2019	13:30
5	CCV1	1	Water	NA	NA	20.2	2.01	02/01/2019	13:31
6	K1332-01	1	Water	NA	NA	13.4	4.73	02/01/2019	13:32
7	K1332-01DUP	1	Water	NA	NA	13.5	4.74	02/01/2019	13:33
8	CCV2	1	Water	NA	NA	20.2	12.01	02/01/2019	13:35

WorkList Name : PH K1332

WorkList ID : 121831

Date : 2/1/2019 12:15:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/30/2019	Water	K1332-01-A	pH	Cool 4 deg C	STER14	L33	MINI-CELL	01/30/2019	SM 4500-H B

Date/Time 02-01-19 12:20 PM
Received by: SD
Relinquished by: [Signature]

Date/Time 02-01-19 11:10 AM
Received by: [Signature]
Relinquished by: JB