



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/4/2019

OVENTEMP IN Celsius(°C): 107  
Time IN: 15:15  
In Date: 02/01/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:36  
Out Date: 02/02/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: % SOLIDS OVEN

QC:LB100595

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1316-01 | MGAET7           | 1     | 1.14              | 9.52                          | 7.98                              | 81.6    |
| K1316-02 | MGAET8           | 2     | 1.18              | 9.96                          | 8.44                              | 82.7    |
| K1316-03 | MGAET9           | 3     | 1.13              | 9.72                          | 9.4                               | 96.3    |
| K1316-04 | MGAET9D          | 4     | 1.13              | 9.72                          | 9.4                               | 96.3    |
| K1316-05 | MGAET9S          | 5     | 1.13              | 9.72                          | 9.4                               | 96.3    |
| K1316-06 | MGAEW0           | 6     | 1.18              | 9.93                          | 8.25                              | 80.8    |
| K1316-07 | MGAEW1           | 7     | 1.14              | 9.91                          | 8.29                              | 81.5    |
| K1316-08 | MGAEW2           | 8     | 1.13              | 9.76                          | 8.19                              | 81.8    |
| K1316-09 | MGAEW3           | 9     | 1.16              | 9.61                          | 7.77                              | 78.2    |
| K1316-10 | MGAEW4           | 10    | 1.17              | 9.79                          | 8.12                              | 80.6    |
| K1316-11 | MGAEW5           | 11    | 1.15              | 9.96                          | 8.00                              | 77.8    |
| K1316-12 | MGAEW6           | 12    | 1.16              | 9.63                          | 7.92                              | 79.8    |
| K1316-13 | MGAEW7           | 13    | 1.18              | 9.75                          | 8.06                              | 80.3    |
| K1316-14 | MGAEW8           | 14    | 1.14              | 9.89                          | 7.52                              | 72.9    |
| K1316-15 | MGAEW9           | 15    | 1.17              | 9.7                           | 8.1                               | 81.2    |
| K1316-16 | MGAEX0           | 16    | 1.18              | 9.97                          | 7.86                              | 76.0    |
| K1316-17 | MGAEX1           | 17    | 1.18              | 9.97                          | 8.06                              | 78.3    |
| K1316-18 | MGAEX2           | 18    | 1.14              | 9.94                          | 7.74                              | 75.0    |
| K1316-19 | MGAEX3           | 19    | 1.19              | 9.92                          | 8.98                              | 89.2    |
| K1316-20 | MGAEX4           | 20    | 1.14              | 9.8                           | 8.4                               | 83.8    |
| K1316-21 | MGAEX5           | 21    | 1.17              | 9.67                          | 8.27                              | 83.5    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$