



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB100599

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include various calibration standards and reagents like hexachrome, hexavalent chromium, and HNO3 Hex-Chrome.

Intercept: 0.0005 Slope: 0.7711 Regression: 0.999981

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB100599

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.25	0.000	0.391	0.391	0.506	02/01/2019	14:03
2	ICB		1	100	100		2.30	0.000	0.000	0.000	-0.001	02/01/2019	14:04
3	CCV1	0.5	1	100	100		2.11	0.000	0.391	0.391	0.506	02/01/2019	14:04
4	CCB1		1	100	100		1.86	0.000	0.000	0.000	-0.001	02/01/2019	14:05
5	PB116769BL		1	2.53	100	7.13	1.74	0.000	0.000	0.000	-0.001	02/01/2019	14:05
6	PB116769BS	20	1	2.50	100	7.56	1.70	0.000	0.366	0.366	0.474	02/01/2019	14:06
7	K1011-13		1	2.55	100	7.80	2.33	0.010	0.010	0.000	-0.001	02/01/2019	14:06
8	K1011-15		1	2.54	100	7.67	2.29	0.013	0.013	0.000	-0.001	02/01/2019	14:07
9	K1297-01		1	2.53	100	7.12	1.59	0.108	0.108	0.000	-0.001	02/01/2019	14:07
10	K1297-01DUP		1	2.50	100	7.31	1.63	0.109	0.109	0.000	-0.001	02/01/2019	14:08
11	K1297-02MSPre	40	2	2.52	100	7.25	1.94	0.114	0.190	0.076	0.098	02/01/2019	14:08
12	K1297-02MS2Ins	1284	40	2.51	100	7.69	2.09	0.105	0.624	0.519	0.672	02/01/2019	14:09
13	K1297-02MS3Post	40	2	2.55	100	7.44	2.40	0.116	0.208	0.092	0.119	02/01/2019	14:09
14	K1297-03	40	2	2.54	100	7.03	2.36	0.114	0.190	0.076	0.098	02/01/2019	14:10
15	CCV2	0.5	1	100	100		1.77	0.000	0.392	0.392	0.508	02/01/2019	14:11
16	CCB2		1	100	100		1.92	0.000	0.000	0.000	-0.001	02/01/2019	14:11
17	K1297-04		1	2.53	100	7.66	2.05	0.088	0.088	0.000	-0.001	02/01/2019	14:12
18	K1297-05		1	2.51	100	7.43	2.10	0.070	0.070	0.000	-0.001	02/01/2019	14:12
19	K1297-06		1	2.52	100	7.38	2.19	0.032	0.032	0.000	-0.001	02/01/2019	14:13
20	K1299-01		1	2.54	100	7.21	2.31	0.091	0.091	0.000	-0.001	02/01/2019	14:13
21	K1299-02		1	2.53	100	7.17	1.84	0.074	0.074	0.000	-0.001	02/01/2019	14:14
22	K1299-03		1	2.53	100	7.69	1.66	0.088	0.088	0.000	-0.001	02/01/2019	14:14
23	K1299-04		1	2.55	100	7.72	1.70	0.080	0.080	0.000	-0.001	02/01/2019	14:15
24	K1299-05		1	2.50	100	7.96	2.14	0.070	0.070	0.000	-0.001	02/01/2019	14:15
25	K1299-06		1	2.51	100	7.82	2.27	0.056	0.056	0.000	-0.001	02/01/2019	14:16
26	K1299-07		1	2.52	100	7.55	2.35	0.047	0.047	0.000	-0.001	02/01/2019	14:16
27	CCV3	0.5	1	100	100		2.42	0.000	0.392	0.392	0.508	02/01/2019	14:17
28	CCB3		1	100	100		1.64	0.000	0.000	0.000	-0.001	02/01/2019	14:17