

LB100604


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55437



A= 0.0000e+000

B= 1.7934e-004

C= -4.6157e-002

Rho= 0.9998319

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	166	0.000	166				
0.2	0.200	0.178	-0.022	1251	0.0 %	1251				
2.5	2.500	2.503	0.003	14215	0.0 %	14215				
5.0	5.000	5.013	0.013	28212	0.0 %	28212				
7.5	7.500	7.625	0.125	42776	0.0 %	42776				
10.0	10.000	9.896	-0.104	55437	0.0 %	55437				

LB100604 INSTRUMENT ID :CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Feb 2019 16:26:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	166	-	0.0000	7471B	PPB	01 Feb 2019 16:29:55	S	Std
0.2 - 1	50.2	1251	-	0.2000	7471B	PPB	01 Feb 2019 16:32:13	S	Std
2.5 - 1	52.5	14215	-	2.5000	7471B	PPB	01 Feb 2019 16:34:30	S	Std
5.0 - 1	55.0	28212	-	5.0000	7471B	PPB	01 Feb 2019 16:36:47	S	Std
7.5 - 1	57.5	42776	-	7.5000	7471B	PPB	01 Feb 2019 16:39:03	S	Std
10.0 - 1	510.0	55437	-	10.0000	7471B	PPB	01 Feb 2019 16:41:20	S	Std
ICV91 - 1	ICV91	22144	3.9252	-	7471B	PPB	01 Feb 2019 16:44:35	U	SMPL
ICB91 - 1	ICB91	-41	-0.0535	-	7471B	PPB	01 Feb 2019 16:46:51	U	SMPL
CCV60 - 1	CCV60	28231	5.0169	-	7471B	PPB	01 Feb 2019 16:49:07	U	SMPL
CCB60 - 1	CCB60	-46	-0.0544	-	7471B	PPB	01 Feb 2019 16:51:22	U	SMPL
CRA - 1	CRA	1503	0.2234	-	7471B	PPB	01 Feb 2019 16:53:38	U	SMPL
HighStd - 1	HighStd	55281	9.8681	-	7471B	PPB	01 Feb 2019 16:58:55	U	SMPL
ChkStd - 1	ChkStd	39555	7.0478	-	7471B	PPB	01 Feb 2019 17:01:10	U	SMPL
PB116754BL - 1	PBS	-10	-0.0480	-	7471B	PPB	01 Feb 2019 17:03:28	U	SMPL
PB116754BS - 1	LCSS	19566	3.4629	-	7471B	PPB	01 Feb 2019 17:05:45	U	SMPL
K1011-13 - 1	OR-02-013119-A	1918	0.2978	-	7471B	PPB	01 Feb 2019 17:08:02	U	SMPL
K1011-15 - 1	OR-02-013119-B	1534	0.2290	-	7471B	PPB	01 Feb 2019 17:10:19	U	SMPL
K1297-01 - 1	A0C17-SSI	33330	5.9313	-	7471B	PPB	01 Feb 2019 17:12:36	U	SMPL
K1297-01DUP - 1	A0C17-SSIDUP	33471	5.9566	-	7471B	PPB	01 Feb 2019 17:14:53	U	SMPL
K1297-02 - 1	K1297-01MS	66430	11.8676	-	7471B	PPB	01 Feb 2019 17:17:10	U	SMPL
CCV61 - 1	CCV61	28110	4.9952	-	7471B	PPB	01 Feb 2019 17:19:27	U	SMPL
CCB61 - 1	CCB61	-108	-0.0655	-	7471B	PPB	01 Feb 2019 17:21:45	U	SMPL
K1297-03 - 1	K1297-01MSD	66270	11.8389	-	7471B	PPB	01 Feb 2019 17:24:03	U	SMPL
K1297-04 - 1	DUP-01-30-19	61029	10.8990	-	7471B	PPB	01 Feb 2019 17:26:18	U	SMPL
K1297-05 - 1	A0C17-SS2	0	-0.0462	-	7471B	PPB	01 Feb 2019 17:28:36	U	SMPL
K1297-05 - 1	A0C17-SS2	13276	2.3348	-	7471B	PPB	01 Feb 2019 17:34:01	U	SMPL
K1297-06 - 1	A0C4-SS1	18971	3.3562	-	7471B	PPB	01 Feb 2019 17:36:16	U	SMPL
K1299-01 - 1	0035-COMP-1	133010	23.8083	-	7471B	PPB	01 Feb 2019 17:38:33	U	SMPL
K1299-02 - 1	0036-COMP-2	128058	22.9202	-	7471B	PPB	01 Feb 2019 17:40:50	U	SMPL
K1299-03 - 1	0037-COMP-3	171292	30.6739	-	7471B	PPB	01 Feb 2019 17:43:10	U	SMPL
K1299-04 - 1	0041-FIELD-DUP	117083	20.9519	-	7471B	PPB	01 Feb 2019 17:45:31	U	SMPL
K1299-05 - 1	0038-AMND-COMP-1	150615	26.9656	-	7471B	PPB	01 Feb 2019 17:47:52	U	SMPL
K1299-06 - 1	0039-AMND-COMP-2	103744	18.5596	-	7471B	PPB	01 Feb 2019 17:50:10	U	SMPL
CCV62 - 1	CCV62	27908	4.9590	-	7471B	PPB	01 Feb 2019 17:52:29	U	SMPL
CCB62 - 1	CCB62	-126	-0.0688	-	7471B	PPB	01 Feb 2019 17:54:48	U	SMPL
K1299-07 - 1	0040-AMND-COMP-3	124804	22.3366	-	7471B	PPB	01 Feb 2019 17:57:04	U	SMPL
K1308-01 - 1	SB-17(34.2-34.8)	-208	-0.0835	-	7471B	PPB	01 Feb 2019 17:59:20	U	SMPL
K1308-02 - 1	SB-18(34.5-35)	199	-0.0105	-	7471B	PPB	01 Feb 2019 18:01:39	U	SMPL
K1297-01LX5 - 1		6462	1.1128	-	7471B	PPB	01 Feb 2019 18:03:55	U	SMPL
K1297-01A - 1		55444	9.8973	-	7471B	PPB	01 Feb 2019 18:06:11	U	SMPL
K1297-01DLX2 - 1		16202	2.8596	-	7471B	PPB	01 Feb 2019 18:08:30	U	SMPL
K1297-01DUPDLX2 - 1		15784	2.7846	-	7471B	PPB	01 Feb 2019 18:10:49	U	SMPL
K1297-01MSDLX2 - 1		31749	5.6478	-	7471B	PPB	01 Feb 2019 18:13:06	U	SMPL
K1297-01MSDDLX2 - 1		32089	5.7088	-	7471B	PPB	01 Feb 2019 18:15:24	U	SMPL
CCV63 - 1	CCV63	28290	5.0275	-	7471B	PPB	01 Feb 2019 18:26:53	U	SMPL
CCB63 - 1	CCB63	22	-0.0422	-	7471B	PPB	01 Feb 2019 18:29:08	U	SMPL
K1297-04DLX2 - 1		30418	5.4091	-	7471B	PPB	01 Feb 2019 18:31:25	U	SMPL
K1299-01DLX4 - 1		33749	6.0065	-	7471B	PPB	01 Feb 2019 18:33:41	U	SMPL
K1299-02DLX4 - 1		31999	5.6926	-	7471B	PPB	01 Feb 2019 18:35:58	U	SMPL
K1299-03DLX5 - 1		36049	6.4190	-	7471B	PPB	01 Feb 2019 18:38:15	U	SMPL
K1299-04DLX4 - 1		29145	5.1808	-	7471B	PPB	01 Feb 2019 18:40:31	U	SMPL
K1299-05DLX5 - 1		30820	5.4812	-	7471B	PPB	01 Feb 2019 18:42:49	U	SMPL
K1299-06DLX4 - 1		25392	4.5077	-	7471B	PPB	01 Feb 2019 18:45:06	U	SMPL
K1299-07DLX4 - 1		30716	5.4625	-	7471B	PPB	01 Feb 2019 18:47:24	U	SMPL
CCV64 - 1	CCV64	27449	4.8766	-	7471B	PPB	01 Feb 2019 18:49:42	U	SMPL
CCB64 - 1	CCB64	-18	-0.0494	-	7471B	PPB	01 Feb 2019 18:51:59	U	SMPL