



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB100605
BalanceID: WC SC-4

Analyst By : Jignesh
Supervisor Review By : apatel
Slope : 98.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/01/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	02/01/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/01/2019	17:12
4	ICV	1	Water	NA	NA	20.1	7.00	02/01/2019	17:20
5	CCV1	1	Water	NA	NA	20.1	2.01	02/01/2019	17:21
6	K1312-02	1	Solid	20.02	20	19.4	5.42	02/01/2019	17:22
7	K1312-02DUP	1	Solid	20.03	20	19.3	5.43	02/01/2019	17:24
8	K1360-02	1	Solid	20.04	20	18.9	6.46	02/01/2019	17:25
9	K1360-04	1	Solid	20.02	20	20.0	6.05	02/01/2019	17:26
10	CCV2	1	Water	NA	NA	20.1	12.01	02/01/2019	17:30

29 100605

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrsivity k1312 WorkList ID : 121852 Date : 2/1/2019 4:48:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/01/2019	Solid	K1312-02	Corrosivity	Cool 4 deg C	PSEG01	K31	SOIL-PILE	02/01/2019	9045D
02/01/2019	Solid	K1360-02	Corrosivity	Cool 4 deg C	PSEG01	K23	RBR-200053	02/01/2019	9045D
02/01/2019	Solid	K1360-04	Corrosivity	Cool 4 deg C	PSEG01	K23	RBR-200034	02/01/2019	9045D

Date/Time 02-01-19 6:50 PM
 Received by: JG
 Relinquished by: CA

Date/Time 02-01-19 6:15 PM
 Received by: CA
 Relinquished by: JG