



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100606

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/01/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	02/01/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/01/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.01	02/01/2019	17:14
5	CCV1	1	Water	NA	NA	20.1	2.01	02/01/2019	17:15
6	K1359-04	1	Water	NA	NA	11.9	6.75	02/01/2019	17:16
7	K1359-04DUP	1	Water	NA	NA	11.8	6.76	02/01/2019	17:17
8	CCV2	1	Water	NA	NA	20.1	12.01	02/01/2019	17:18

2B100606

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrsoivity k1359W

WorkList ID : 121853

Date : 2/1/2019 4:49:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/01/2019	Water	K1359-04	Corrosivity	Cool 4 deg C	PSEG01	K31	COMP-3	02/01/2019	9040C

Date/Time 02-01-19 4:50 PM
 Received by: Jo
 Relinquished by: LG

Date/Time 02-01-19 6:15 PM
 Received by: LG
 Relinquished by: Jo