



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/4/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 15:10  
In Date: 02/03/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:37  
Out Date: 02/04/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: % SOLIDS OVEN

QC:LB100615

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1324-01 | MGAF90           | 1     | 1.18              | 9.75                          | 8.14                              | 81.2    |
| K1324-02 | MGAF91           | 2     | 1.17              | 9.94                          | 7.63                              | 73.7    |
| K1324-03 | MGAF92           | 3     | 1.12              | 9.93                          | 7.92                              | 77.2    |
| K1324-04 | MGAF93           | 4     | 1.18              | 9.77                          | 8.76                              | 88.2    |
| K1324-05 | MGAF94           | 5     | 1.1               | 9.94                          | 7.87                              | 76.6    |
| K1324-06 | MGAF95           | 6     | 1.18              | 9.98                          | 8.03                              | 77.8    |
| K1324-07 | MGAF96           | 7     | 1.1               | 9.91                          | 8.66                              | 85.8    |
| K1324-08 | MGAF97           | 8     | 1.19              | 9.96                          | 8.55                              | 83.9    |
| K1324-09 | MGAF98           | 9     | 1.11              | 9.9                           | 8.09                              | 79.4    |
| K1324-10 | MGAF99           | 10    | 1.17              | 9.88                          | 7.86                              | 76.8    |
| K1324-11 | MGAF99D          | 11    | 1.17              | 9.88                          | 7.86                              | 76.8    |
| K1324-12 | MGAF99S          | 12    | 1.17              | 9.88                          | 7.86                              | 76.8    |
| K1324-13 | MGAF9A0          | 13    | 1.1               | 9.86                          | 7.92                              | 77.9    |
| K1324-14 | MGAF9A1          | 14    | 1.13              | 9.53                          | 7.48                              | 75.6    |
| K1324-15 | MGAF9A2          | 15    | 1.19              | 9.97                          | 6.92                              | 65.3    |
| K1324-16 | MGAF9A3          | 16    | 1.15              | 9.51                          | 8.03                              | 82.3    |
| K1324-17 | MGAF9A4          | 17    | 1.19              | 9.77                          | 8.31                              | 83.0    |
| K1324-18 | MGAF9A5          | 18    | 1.13              | 9.94                          | 8.22                              | 80.5    |
| K1324-19 | MGAF9A6          | 19    | 1.13              | 9.95                          | 8.39                              | 82.3    |
| K1324-20 | MGAF9A7          | 20    | 1.19              | 9.81                          | 7.8                               | 76.7    |
| K1324-21 | MGAF9A8          | 21    | 1.1               | 9.67                          | 7.52                              | 74.9    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$