



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/4/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 15:30
In Date: 02/03/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:43
Out Date: 02/04/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-1
Thermometer ID: % SOLIDS OVEN

QC:LB100616

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1325-01	MGAFB9	1	1.13	9.57	7.68	77.6
K1325-02	MGAFB0	2	1.19	9.97	8.86	87.4
K1325-03	MGAFB1	3	1.17	9.68	8.00	80.3
K1325-04	MGAFB2	4	1.17	9.98	8.74	85.9
K1325-05	MGAFB3	5	1.15	9.95	9.22	91.7
K1325-06	MGAFB4	6	1.16	9.65	8.07	81.4
K1325-07	MGAFB5	7	1.16	9.68	7.98	80.0
K1325-08	MGAFB6	8	1.1	9.54	8.93	92.8
K1325-09	MGAFB7	9	1.17	9.91	9.51	95.4
K1325-10	MGAFB8	10	1.15	9.55	7.9	80.4
K1325-11	MGAFB9	11	1.18	9.92	9.32	93.1
K1325-12	MGAFB9D	12	1.18	9.92	9.32	93.1
K1325-13	MGAFB9S	13	1.18	9.92	9.32	93.1
K1325-14	MGAFB0	14	1.18	9.97	8.7	85.6
K1325-15	MGAFB1	15	1.13	9.97	8.84	87.2
K1325-16	MGAFB2	16	1.13	9.65	7.55	75.4
K1325-17	MGAFB3	17	1.19	9.98	8.32	81.1
K1325-18	MGAFB4	18	1.17	9.94	8.41	82.6
K1325-19	MGAFB5	19	1.18	9.67	7.7	76.8
K1325-20	MGAFB6	20	1.18	9.9	8.16	80.0
K1325-21	MGAFB7	21	1.14	9.73	7.89	78.6
K1325-22	MGAFB3	22	1.13	9.55	7.96	81.1

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$