



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/4/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 15:50  
In Date: 02/03/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:47  
Out Date: 02/04/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: % SOLIDS OVEN

QC:LB100617

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K1326-01	MGAF C8	1	1.18	9.7	7.44	73.5
K1326-02	MGAF C9	2	1.16	9.98	8.2	79.8
K1326-03	MGAF D0	3	1.1	9.53	7.95	81.3
K1326-04	MGAF D1	4	1.18	9.65	8.00	80.5
K1326-05	MGAF D2	5	1.15	9.96	8.41	82.4
K1326-06	MGAF D3	6	1.13	9.55	8.76	90.6
K1326-07	MGAF D4	7	1.14	9.94	9.14	90.9
K1326-08	MGAF D5	8	1.14	9.81	8.4	83.7
K1326-09	MGAF D6	9	1.19	9.97	8.35	81.5
K1326-10	MGAF D7	10	1.15	9.96	8.51	83.5
K1326-11	MGAF D8	11	1.17	9.59	8.34	85.2
K1326-12	MGAF D9	12	1.13	9.94	8.12	79.3
K1326-13	MGAF D9D	13	1.13	9.94	8.12	79.3
K1326-14	MGAF D9S	14	1.13	9.94	8.12	79.3
K1326-15	MGAF E0	15	1.19	9.96	8.13	79.1
K1326-16	MGAF E1	16	1.2	9.98	8.46	82.7
K1326-17	MGAF E2	17	1.11	9.61	7.77	78.4
K1326-18	MGAF E3	18	1.17	9.78	7.4	72.4
K1326-19	MGAF E4	19	1.16	9.78	8.35	83.4
K1326-20	MGAF E5	20	1.13	9.56	8.06	82.2
K1326-21	MGAF E6	21	1.17	9.85	7.78	76.2
K1326-22	MGAF N4	22	1.18	9.94	8.31	81.4

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$