



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/5/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 15:00  
In Date: 02/04/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:32  
Out Date: 02/05/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: % SOLIDS OVEN

QC:LB100625

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1331-01 | MGAF14           | 1     | 1.13              | 9.62                          | 7.96                              | 80.4    |
| K1331-02 | MGAF15           | 2     | 1.15              | 9.96                          | 8.2                               | 80.0    |
| K1331-03 | MGAF16           | 3     | 1.17              | 9.92                          | 8.32                              | 81.7    |
| K1331-04 | MGAF17           | 4     | 1.18              | 9.96                          | 8.25                              | 80.5    |
| K1331-05 | MGAF18           | 5     | 1.15              | 9.93                          | 8.42                              | 82.8    |
| K1331-06 | MGAF19           | 6     | 1.17              | 9.76                          | 8.23                              | 82.2    |
| K1331-07 | MGAFM0           | 7     | 1.17              | 9.5                           | 7.88                              | 80.6    |
| K1331-08 | MGAFM1           | 8     | 1.14              | 9.98                          | 8.42                              | 82.4    |
| K1331-09 | MGAFM2           | 9     | 1.17              | 9.97                          | 8.45                              | 82.7    |
| K1331-10 | MGAFM3           | 10    | 1.16              | 9.67                          | 7.52                              | 74.7    |
| K1331-11 | MGAFM4           | 11    | 1.17              | 9.7                           | 7.35                              | 72.5    |
| K1331-12 | MGAFM5           | 12    | 1.16              | 9.92                          | 8.71                              | 86.2    |
| K1331-13 | MGAFM6           | 13    | 1.14              | 9.57                          | 7.55                              | 76.0    |
| K1331-14 | MGAFM7           | 14    | 1.1               | 9.96                          | 7.95                              | 77.3    |
| K1331-15 | MGAFM8           | 15    | 1.15              | 9.94                          | 7.73                              | 74.9    |
| K1331-16 | MGAFM9           | 16    | 1.14              | 9.54                          | 8.15                              | 83.5    |
| K1331-17 | MGAFM9D          | 17    | 1.14              | 9.54                          | 8.15                              | 83.5    |
| K1331-18 | MGAFM9S          | 18    | 1.14              | 9.54                          | 8.15                              | 83.5    |
| K1331-19 | MGAFN0           | 19    | 1.19              | 9.78                          | 8.25                              | 82.2    |
| K1331-20 | MGAFN1           | 20    | 1.17              | 9.94                          | 9.09                              | 90.3    |
| K1331-21 | MGAFN2           | 21    | 1.16              | 9.77                          | 7.51                              | 73.8    |
| K1331-22 | MGAFN8           | 22    | 1.17              | 9.91                          | 7.7                               | 74.7    |
| K1331-23 | MGAFN9           | 23    | 1.00              | 2.00                          | 2.00                              | 100.0   |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$