

DB106633

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/4/2019 11:54

Test: Total CN

mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.664	0.0	0.093	
ICB1	-0.472	0.0	0.004	
CCV1	238.879	0.0	0.227	
CCB1	-0.899	0.0	0.004	
PB116771BL	-0.774	0.0	0.004	
PB116771BS	96.333	0.0	0.094	
MIDPB116771	240.110	0.0	0.228	
K1239-02	9.411	0.0	0.013	
K1239-02DUP	9.353	0.0	0.013	
K1239-02MS	42.436	0.0	0.044	
K1239-02MSD	45.401	0.0	0.047	
CCV2	254.831	0.0	0.241	
CCB2	-0.827	0.0	0.004	
N	13			
Mean	79.111			
SD	100.2391			
CV%	126.71			

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 04 11:46:54 2019

Mon Feb 04 11:51:16 2019



Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.7113	µg/l	2/4/2019 8:48:02	
5.OPPBCN	A	Total CN	P	4.2759	µg/l	2/4/2019 8:48:03	
10PPBCN	A	Total CN	P	9.6024	µg/l	2/4/2019 8:48:04	
50PPBCN	A	Total CN	P	49.4446	µg/l	2/4/2019 8:48:05	
100PPBCN	A	Total CN	P	100.7902	µg/l	2/4/2019 8:48:06	
250PPBCN	A	Total CN	P	253.3708	µg/l	2/4/2019 8:48:07	
500PPBCN	A	Total CN	P	498.2273	µg/l	2/4/2019 8:48:08	
ICV1	S	Total CN	P	94.6638	µg/l	2/4/2019 11:46:54	
ICB1	S	Total CN	P	-0.4725	µg/l	2/4/2019 11:46:55	
CCV1	S	Total CN	P	238.8786	µg/l	2/4/2019 11:46:56	
CCB1	S	Total CN	P	-0.8988	µg/l	2/4/2019 11:46:57	
PB116771BL	S	Total CN	P	-0.7737	µg/l	2/4/2019 11:46:58	
PB116771BS	S	Total CN	P	96.3331	µg/l	2/4/2019 11:46:59	
MIDPB116771	S	Total CN	P	240.1099	µg/l	2/4/2019 11:47:00	
K1239-02	S	Total CN	P	9.4111	µg/l	2/4/2019 11:47:01	
K1239-02DUP	S	Total CN	P	9.3531	µg/l	2/4/2019 11:47:02	
K1239-02MS	S	Total CN	P	42.436	µg/l	2/4/2019 11:47:03	
K1239-02MSD	S	Total CN	P	45.401	µg/l	2/4/2019 11:47:04	
CCV2	S	Total CN	P	254.8307	µg/l	2/4/2019 11:51:15	
CCB2	S	Total CN	P	-0.8269	µg/l	2/4/2019 11:51:16	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/4/2019 9:01


Test Total CN

Accepted 2/4/2019 9:00

Factor 1076
Bias 0.005

Coeff. of det. 0.999919

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.7113	0.0000	
2	5.0PPBCN	0.009	4.2759	5.0000	
3	10PPBCN	0.014	9.6024	10.0000	
4	50PPBCN	0.051	49.4446	50.0000	
5	100PPBCN	0.098	100.7902	100.0000	
6	250PPBCN	0.240	253.3708	250.0000	
7	500PPBCN	0.467	498.2273	500.0000	