


LB100634

7470A

INSTRUMENT ID : CV1

 μ Abs. :

41037



Linear

A= 0.0000e+000

B= 2.4415e-004

C= -1.1041e-002

Rho= 0.9998231

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	36	0.000	36				
0.2	0.200	0.174	-0.026	756	0.0 %	756				
2.5	2.500	2.484	-0.016	10221	0.0 %	10221				
5.0	5.000	5.135	0.135	21077	0.0 %	21077				
7.5	7.500	7.401	-0.099	30361	0.0 %	30361				
10.0	10.000	10.008	0.008	41037	0.0 %	41037				

LB100634 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 04 Feb 2019 12:07:26				
Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50-0	36	-	0.00007470A	PPB	04 Feb 2019 12:20:29	S	Std
0.2 - 1	50-2	756	-	0.20007470A	PPB	04 Feb 2019 12:22:37	S	Std
2.5 - 1	52-5	10221	-	2.50007470A	PPB	04 Feb 2019 12:24:45	S	Std
5.0 - 1	55-0	21077	-	5.00007470A	PPB	04 Feb 2019 12:26:53	S	Std
7.5 - 1	57-5	30361	-	7.50007470A	PPB	04 Feb 2019 12:32:28	S	Std
10.0 - 1	510-0	41037	-	10.00007470A	PPB	04 Feb 2019 12:39:39	S	Std
ICV92 - 1	ICV92	14810	3.6048	-7470A	PPB	04 Feb 2019 12:42:44	U	SMPL
ICB92 - 1	ICB92	-23	-0.0167	-7470A	PPB	04 Feb 2019 12:44:50	U	SMPL
CCV65 - 1	CCV65	21028	5.1229	-7470A	PPB	04 Feb 2019 12:46:56	U	SMPL
CCB65 - 1	CCB65	-19	-0.0157	-7470A	PPB	04 Feb 2019 12:49:02	U	SMPL
CRA - 1	CRA	827	0.1909	-7470A	PPB	04 Feb 2019 12:51:08	U	SMPL
HighStd - 1	HighStd	40846	9.9614	-7470A	PPB	04 Feb 2019 12:53:14	U	SMPL
ChkStd - 1	ChkStd	28642	6.9818	-7470A	PPB	04 Feb 2019 12:55:20	U	SMPL
PB116773BL - 1	PBW	-29	-0.0181	-7470A	PPB	04 Feb 2019 12:57:27	U	SMPL
PB116773BS - 1	LCSW	15501	3.7735	-7470A	PPB	04 Feb 2019 12:59:34	U	SMPL
K1299-08 - 1	SPLP-LEACH-AMND-COMP-1	149	0.0253	-7470A	PPB	04 Feb 2019 13:01:41	U	SMPL
K1299-08DUP - 1	SPLP-LEACH-AMND-COMP-1DUP	64	0.0046	-7470A	PPB	04 Feb 2019 13:03:48	U	SMPL
K1299-08MS - 1	SPLP-LEACH-AMND-COMP-1MS	15500	3.7732	-7470A	PPB	04 Feb 2019 13:05:56	U	SMPL
K1299-08MSD - 1	SPLP-LEACH-AMND-COMP-1MSD	15368	3.7410	-7470A	PPB	04 Feb 2019 13:08:04	U	SMPL
K1299-09 - 1	SPLP-LEACH-AMND-COMP-2	53	0.0019	-7470A	PPB	04 Feb 2019 13:10:11	U	SMPL
CCV66 - 1	CCV66	21143	5.1510	-7470A	PPB	04 Feb 2019 13:12:17	U	SMPL
CCB66 - 1	CCB66	-24	-0.0169	-7470A	PPB	04 Feb 2019 13:14:24	U	SMPL
K1299-10 - 1	SPLP-LEACH-AMND-COMP-3	135	0.0219	-7470A	PPB	04 Feb 2019 13:16:30	U	SMPL
PB116775BL - 1	PBW	33	-0.0030	-7470A	PPB	04 Feb 2019 13:18:37	U	SMPL
PB116775BS - 1	LCSW	14658	3.5677	-7470A	PPB	04 Feb 2019 13:20:43	U	SMPL
K1011-14 - 1	OR-02-013119-A	10	-0.0086	-7470A	PPB	04 Feb 2019 13:22:50	U	SMPL
K1011-16 - 1	OR-02-013119-B	28	-0.0042	-7470A	PPB	04 Feb 2019 13:24:56	U	SMPL
K1292-02 - 1	WC-1	18	-0.0066	-7470A	PPB	04 Feb 2019 13:27:04	U	SMPL
K1292-02DUP - 1	WC-1DUP	44	-0.0003	-7470A	PPB	04 Feb 2019 13:29:11	U	SMPL
K1292-02MS - 1	WC-1MS	14914	3.6302	-7470A	PPB	04 Feb 2019 13:31:19	U	SMPL
K1292-02MSD - 1	WC-1MSD	14934	3.6350	-7470A	PPB	04 Feb 2019 13:33:26	U	SMPL
K1303-01 - 1	131-A-E	-17	-0.0152	-7470A	PPB	04 Feb 2019 13:35:33	U	SMPL
CCV67 - 1	CCV67	21130	5.1478	-7470A	PPB	04 Feb 2019 13:37:39	U	SMPL
CCB67 - 1	CCB67	-50	-0.0232	-7470A	PPB	04 Feb 2019 13:39:46	U	SMPL
K1303-02 - 1	131-F-J	40	-0.0013	-7470A	PPB	04 Feb 2019 13:41:52	U	SMPL
K1303-03 - 1	131-K-O	56	0.0026	-7470A	PPB	04 Feb 2019 13:43:59	U	SMPL
K1303-04 - 1	131-P-T	32	-0.0032	-7470A	PPB	04 Feb 2019 13:46:05	U	SMPL
K1303-05 - 1	131-U-X	28	-0.0042	-7470A	PPB	04 Feb 2019 13:48:11	U	SMPL
K1303-06 - 1	131-Y-Z	31	-0.0035	-7470A	PPB	04 Feb 2019 13:50:18	U	SMPL
PB116814BL - 1	PBW	13	-0.0079	-7470A	PPB	04 Feb 2019 13:52:26	U	SMPL
PB116814BS - 1	LCSW	15629	3.8047	-7470A	PPB	04 Feb 2019 13:54:33	U	SMPL
K1296-02 - 1	BP-TT-FB01-20190129	9	-0.0088	-7470A	PPB	04 Feb 2019 13:56:41	U	SMPL
K1296-02DUP - 1	BP-TT-FB01-20190129DUP	19	-0.0064	-7470A	PPB	04 Feb 2019 13:58:49	U	SMPL
K1296-02MS - 1	BP-TT-FB01-20190129MS	15145	3.6866	-7470A	PPB	04 Feb 2019 14:00:55	U	SMPL
CCV68 - 1	CCV68	20516	4.9979	-7470A	PPB	04 Feb 2019 14:03:02	U	SMPL
CCB68 - 1	CCB68	-29	-0.0181	-7470A	PPB	04 Feb 2019 14:05:08	U	SMPL
K1296-02MSD - 1	BP-TT-FB01-20190129MSD	15372	3.7420	-7470A	PPB	04 Feb 2019 14:07:15	U	SMPL
K1296-03 - 1	BP-TT-MW202S-20190129	20	-0.0062	-7470A	PPB	04 Feb 2019 14:09:21	U	SMPL
K1296-06 - 1	BP-TT-RB01-20190129	13	-0.0079	-7470A	PPB	04 Feb 2019 14:11:28	U	SMPL
K1296-07 - 1	BP-TT-MW202I-20190129	6	-0.0096	-7470A	PPB	04 Feb 2019 14:13:34	U	SMPL
K1296-08 - 1	BP-TT-DUP01-20190129	43	-0.0005	-7470A	PPB	04 Feb 2019 14:15:41	U	SMPL
K1296-09 - 1	BP-TT-MW203I-20190130	34	-0.0027	-7470A	PPB	04 Feb 2019 14:17:48	U	SMPL
K1297-07 - 1	FB-01-30-19	77	0.0078	-7470A	PPB	04 Feb 2019 14:19:56	U	SMPL
K1299-11 - 1	0034-FIELD-BLANK	23	-0.0054	-7470A	PPB	04 Feb 2019 14:22:03	U	SMPL
K1302-01 - 1	EFFLUENT-1	211	0.0405	-7470A	PPB	04 Feb 2019 14:24:12	U	SMPL
PB116773TB - 1	PB116740TB	23	-0.0054	-7470A	PPB	04 Feb 2019 14:26:18	U	SMPL
CCV69 - 1	CCV69	20294	4.9437	-7470A	PPB	04 Feb 2019 14:28:24	U	SMPL
CCB69 - 1	CCB69	-14	-0.0145	-7470A	PPB	04 Feb 2019 14:30:31	U	SMPL
PB116775TB - 1	PB116732TB	53	0.0019	-7470A	PPB	04 Feb 2019 14:32:38	U	SMPL
K1299-08LX5 - 1		36	-0.0023	-7470A	PPB	04 Feb 2019 14:34:44	U	SMPL
K1299-08A - 1		16316	3.9725	-7470A	PPB	04 Feb 2019 14:36:50	U	SMPL
K1292-02LX5 - 1		-33	-0.0191	-7470A	PPB	04 Feb 2019 14:38:57	U	SMPL
K1292-02A - 1		16322	3.9739	-7470A	PPB	04 Feb 2019 14:41:03	U	SMPL
K1296-02LX5 - 1		-24	-0.0169	-7470A	PPB	04 Feb 2019 14:43:10	U	SMPL
K1296-02A - 1		16303	3.9693	-7470A	PPB	04 Feb 2019 14:45:18	U	SMPL
CCV70 - 1	CCV70	20803	5.0679	-7470A	PPB	04 Feb 2019 14:47:26	U	SMPL
CCB70 - 1	CCB70	-28	-0.0179	-7470A	PPB	04 Feb 2019 14:49:34	U	SMPL