

2B100636

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/4/2019 12:25

Test: *Reactive*
~~Total~~ CN

ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	90.498	0.0	0.089	
ICB1	-0.695	0.0	0.004	
CCV1	234.123	0.0	0.222	
CCB1	-0.859	0.0	0.004	
PB116765BL	-0.902	0.0	0.004	
K1011-14	-1.044	0.0	0.004	
K1011-14DUP	-0.835	0.0	0.004	
K1011-16	-0.902	0.0	0.004	
K1292-02	-0.694	0.0	0.004	
CCV2	240.413	0.0	0.228	
CCB2	-0.885	0.0	0.004	

N	11
Mean	50.747
SD	96.1666
CV%	189.50

02/04/19
AS

ICV1
ICB1

42

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 04 12:23:53 2019

Mon Feb 04 12:24:03 2019

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.7113	µg/l	2/4/2019 8:48:02	
5.OPPBCN	A	Reactive CN	P	4.2759	µg/l	2/4/2019 8:48:03	
10PPBCN	A	Reactive CN	P	9.6024	µg/l	2/4/2019 8:48:04	
50PPBCN	A	Reactive CN	P	49.4446	µg/l	2/4/2019 8:48:05	
100PPBCN	A	Reactive CN	P	100.7902	µg/l	2/4/2019 8:48:06	
250PPBCN	A	Reactive CN	P	253.3708	µg/l	2/4/2019 8:48:07	
500PPBCN	A	Reactive CN	P	498.2273	µg/l	2/4/2019 8:48:08	
ICV1	S	Reactive CN	P	90.4984	µg/l	2/4/2019 12:23:53	
ICB1	S	Reactive CN	P	-0.6954	µg/l	2/4/2019 12:23:54	
CCV1	S	Reactive CN	P	234.123	µg/l	2/4/2019 12:23:55	
CCB1	S	Reactive CN	P	-0.8586	µg/l	2/4/2019 12:23:56	
PB116765BL	S	Reactive CN	P	-0.9021	µg/l	2/4/2019 12:23:57	
K1011-14	S	Reactive CN	P	-1.0444	µg/l	2/4/2019 12:23:58	
K1011-14DUP	S	Reactive CN	P	-0.8348	µg/l	2/4/2019 12:23:59	
K1011-16	S	Reactive CN	P	-0.9025	µg/l	2/4/2019 12:24:00	
K1292-02	S	Reactive CN	P	-0.6939	µg/l	2/4/2019 12:24:01	
CCV2	S	Reactive CN	P	240.4132	µg/l	2/4/2019 12:24:02	
CCB2	S	Reactive CN	P	-0.8854	µg/l	2/4/2019 12:24:03	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/4/2019 12:38


Test Total CN

Accepted 2/4/2019 9:00

Factor 1076
Bias 0.005

Coeff. of det. 0.999919

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.7113	0.0000	
2	5.0PPBCN	0.009	4.2759	5.0000	
3	10PPBCN	0.014	9.6024	10.0000	
4	50PPBCN	0.051	49.4446	50.0000	
5	100PPBCN	0.098	100.7902	100.0000	
6	250PPBCN	0.240	253.3708	250.0000	
7	500PPBCN	0.467	498.2273	500.0000	