



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100642

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.0       | 80.6      | 80.9            | 02/04/2019 | 12:50     |
| 2   | K1334-03    |               | 1  | 8                 | >100.0     | >212.0    | >212.0          | 02/04/2019 | 13:15     |
| 3   | K1334-03DUP |               | 1  | 8                 | >100.0     | >212.0    | >212.0          | 02/04/2019 | 13:40     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

# WORKLIST(Hardcopy Internal Chain)

LB100642

WorkList Name : FP2119-1

WorkList ID : 121856

Date : 2/1/2019 5:02:30 PM

| Due Date   | Matrix | Sample     | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|------------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 02/06/2019 | Solid  | K1334-03 B | Flash Point | Cool 4 deg C | PSEG01   | K31              | 50151-TOTE      | 02/01/2019   | 1010A  |

Date/Time 02/04/19 12:15 PM  
 Received by: KS  
 Relinquished by: CP

Date/Time 02/04/19 13:45  
 Received by: CP  
 Relinquished by: KS