



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB100643

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	02/04/2019	14:02
2	K1359-01		1	7	>100.0	>212.0	>212.0	02/04/2019	14:32
3	K1359-01DUP		1	7	>100.0	>212.0	>212.0	02/04/2019	14:50
4	K1359-02		1	9	>100.0	>212.0	>212.0	02/04/2019	15:20
5	K1359-04		1	8	>100.0	>212.0	>212.0	02/04/2019	15:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB100643

WorkList Name : FP2419

WorkList ID : 121893

Date : 2/4/2019 9:40:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2019	Water	K1359-01 <u>D</u>	Flash Point	Cool 4 deg C	PSEG01	K31	COMP-1	02/01/2019	1010A
02/06/2019	Water	K1359-02 <u>F</u>	Flash Point	Cool 4 deg C	PSEG01	K31	COMP-2	02/01/2019	1010A
02/06/2019	Water	K1359-04 <u>C</u>	Flash Point	Cool 4 deg C	PSEG01	K31	COMP-3	02/01/2019	1010A

Date/Time 02/04/19 12:15 PM
 Received by: KS
 Relinquished by: CP

Date/Time 02/04/19 15:55
 Received by: CP
 Relinquished by: KS