

LB100648

245.1

INSTRUMENT ID : CV1

 μ Abs. :

38025

Linear


A= 0.0000e+000

B= 2.6369e-004

C= 8.3614e-003

Rho= 0.9998187

Accept=Accepted



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	22	0.000	22				
0.05	0.050	0.062	0.012	205	0.0 %	205				
0.2	0.200	0.215	0.015	784	0.0 %	784				
2.5	2.500	2.548	0.048	9630	0.0 %	9630				
5.0	5.000	4.830	-0.170	18286	0.0 %	18286				
7.5	7.500	7.545	0.045	28582	0.0 %	28582				
10.0	10.000	10.035	0.035	38025	0.0 %	38025				

LB100648 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Feb 2019 16:02:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	22	-	0.0000	245.1	PPB	04 Feb 2019 16:03:58	S	Std
0.05 - 1	50-05	205	-	0.0500	245.1	PPB	04 Feb 2019 16:06:16	S	Std
0.2 - 1	50-2	784	-	0.2000	245.1	PPB	04 Feb 2019 16:08:33	S	Std
2.5 - 1	52-5	9630	-	2.5000	245.1	PPB	04 Feb 2019 16:10:50	S	Std
5.0 - 1	55-0	18286	-	5.0000	245.1	PPB	04 Feb 2019 16:13:07	S	Std
7.5 - 1	57-5	28582	-	7.5000	245.1	PPB	04 Feb 2019 16:15:23	S	Std
10.0 - 1	510-0	38025	-	10.0000	245.1	PPB	04 Feb 2019 16:17:38	S	Std
ICV93 - 1	ICV93	15178	4.0107	-245.1		PPB	04 Feb 2019 16:23:51	U	SMPL
ICB93 - 1	ICB93	-38	-0.0017	-245.1		PPB	04 Feb 2019 16:26:07	U	SMPL
CCV71 - 1	CCV71	19137	5.0546	-245.1		PPB	04 Feb 2019 16:28:21	U	SMPL
CCB71 - 1	CCB71	-59	-0.0072	-245.1		PPB	04 Feb 2019 16:30:37	U	SMPL
CRA - 1	CRA	184	0.0569	-245.1		PPB	04 Feb 2019 16:32:52	U	SMPL
HighStd - 1	HighStd	38151	10.0684	-245.1		PPB	04 Feb 2019 16:35:07	U	SMPL
ChkStd - 1	ChkStd	26554	7.0104	-245.1		PPB	04 Feb 2019 16:37:22	U	SMPL
PB116778BL - 1	PBW	-36	-0.0011	-245.1		PPB	04 Feb 2019 16:39:39	U	SMPL
PB116778BS - 1	LCSW	15681	4.1433	-245.1		PPB	04 Feb 2019 16:41:56	U	SMPL
K1268-05 - 1	012819-SIDI-E-D-B	19	0.0134	-245.1		PPB	04 Feb 2019 16:44:13	U	SMPL
K1268-05DUP - 1	012819-SIDI-E-D-BDUP	35	0.0176	-245.1		PPB	04 Feb 2019 16:46:30	U	SMPL
K1268-05MS - 1	012819-SIDI-E-D-BMS	16797	4.4376	-245.1		PPB	04 Feb 2019 16:48:47	U	SMPL
K1268-05MSD - 1	012819-SIDI-E-D-BMSD	16904	4.4658	-245.1		PPB	04 Feb 2019 16:51:04	U	SMPL
K1268-06 - 1	012819-SIDI-E-D-M	38	0.0184	-245.1		PPB	04 Feb 2019 16:53:19	U	SMPL
CCV72 - 1	CCV72	19133	5.0536	-245.1		PPB	04 Feb 2019 16:55:34	U	SMPL
CCB72 - 1	CCB72	-44	-0.0032	-245.1		PPB	04 Feb 2019 16:57:49	U	SMPL
K1268-07 - 1	012819-SIDI-E-D-S	45	0.0202	-245.1		PPB	04 Feb 2019 17:00:05	U	SMPL
K1268-08 - 1	012819-SIDI-E-U-B	59	0.0239	-245.1		PPB	04 Feb 2019 17:02:20	U	SMPL
K1268-09 - 1	012819-SIDI-E-U-M	55	0.0229	-245.1		PPB	04 Feb 2019 17:04:35	U	SMPL
K1268-10 - 1	012819-SIDI-E-U-S	55	0.0229	-245.1		PPB	04 Feb 2019 17:06:50	U	SMPL
K1268-11 - 1	012819-TIPC-E-D-M	78	0.0289	-245.1		PPB	04 Feb 2019 17:09:07	U	SMPL
K1268-12 - 1	012819-TIPC-E-U-M	87	0.0313	-245.1		PPB	04 Feb 2019 17:11:23	U	SMPL
K1268-13 - 1	012819-DBRI-F-D-M	70	0.0268	-245.1		PPB	04 Feb 2019 17:13:40	U	SMPL
K1268-14 - 1	012819-DBRI-F-U-M	89	0.0318	-245.1		PPB	04 Feb 2019 17:15:57	U	SMPL
K1290-01 - 1	012919-SIDI-E-D-B	88	0.0316	-245.1		PPB	04 Feb 2019 17:18:14	U	SMPL
K1290-02 - 1	012919-SIDI-E-D-M	101	0.0350	-245.1		PPB	04 Feb 2019 17:20:30	U	SMPL
CCV73 - 1	CCV73	19204	5.0723	-245.1		PPB	04 Feb 2019 17:22:45	U	SMPL
CCB73 - 1	CCB73	-43	-0.0030	-245.1		PPB	04 Feb 2019 17:25:01	U	SMPL
K1290-03 - 1	012919-SIDI-E-D-S	87	0.0313	-245.1		PPB	04 Feb 2019 17:27:16	U	SMPL
K1290-04 - 1	012919-SIDI-E-U-B	22	0.0142	-245.1		PPB	04 Feb 2019 17:33:53	U	SMPL
K1290-05 - 1	012919-SIDI-E-U-M	100	0.0347	-245.1		PPB	04 Feb 2019 17:36:08	U	SMPL
K1290-06 - 1	012919-SIDI-E-U-S	90	0.0321	-245.1		PPB	04 Feb 2019 17:38:23	U	SMPL
K1268-05LX5 - 1		85	0.0308	-245.1		PPB	04 Feb 2019 17:40:39	U	SMPL
K1290-04 - 1	012919-SIDI-E-U-B	162	0.0511	-245.1		PPB	04 Feb 2019 17:29:32	U	SMPL
K1290-05 - 1	012919-SIDI-E-U-M	0	0.0084	-245.1		PPB	04 Feb 2019 17:31:47	U	SMPL
K1268-05A - 1		18545	4.8985	-245.1		PPB	04 Feb 2019 18:47:40	U	SMPL
CCV74 - 1	CCV74	19331	5.1058	-245.1		PPB	04 Feb 2019 18:49:57	U	SMPL
CCB74 - 1	CCB74	-79	-0.0125	-245.1		PPB	04 Feb 2019 18:52:14	U	SMPL