



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100650

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	02/04/2019	16:00
2	CAL2	1	Water	NA	NA	21.1	7.01	02/04/2019	16:01
3	CAL3	1	Water	NA	NA	21.1	10.02	02/04/2019	16:03
4	ICV	1	Water	NA	NA	21.2	6.99	02/04/2019	16:05
5	CCV1	1	Water	NA	NA	21.2	2.02	02/04/2019	16:25
6	K1359-01	1	Water	NA	NA	11.5	5.85	02/04/2019	16:26
7	K1359-01DUP	1	Water	NA	NA	11.6	5.86	02/04/2019	16:27
8	K1359-02	1	Water	NA	NA	11.7	5.90	02/04/2019	16:29
9	CCV2	1	Water	NA	NA	21.1	12.01	02/04/2019	16:30

13100650

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1359

WorkList ID : 121923

Date : 2/4/2019 4:15:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/01/2019	Water	K1359-01-A	pH	Cool 4 deg C	PSEG01	K31	COMP-1	02/01/2019	9040C
02/01/2019	Water	K1359-02-A	pH	Cool 4 deg C	PSEG01	K31	COMP-2	02/01/2019	9040C

Date/Time 02/04/19 6:20 PM
 Received by: Jn
 Relinquished by: CJ

Date/Time 02/04/19 5:00 PM
 Received by: Jn
 Relinquished by: Jn