


LB100653

INSTRUMENT ID : CV1

ISM02.4

Linear



A= 0.0000e+000

B= 2.0615e-004

C= -6.6540e-002

Rho= 0.9999307

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.027	-0.027	193	0.000	193					
0.2	0.200	0.162	-0.038	1109	0.0 %	1109					-20
2.5	2.500	2.561	0.061	12745	0.0 %	12745					-3
5.0	5.000	5.017	0.017	24659	0.0 %	24659					000
7.5	7.500	7.544	0.044	36918	0.0 %	36918					000
10.0	10.000	9.943	-0.057	48555	0.0 %	48555					000

LB100653

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		193	-	0	ISM02.4	PPB	2/5/2019 10:09	Std
0.2		1109	-	0.2	ISM02.4	PPB	2/5/2019 10:11	Std
2.5		12745	-	2.5	ISM02.4	PPB	2/5/2019 10:14	Std
5		24659	-	5	ISM02.4	PPB	2/5/2019 10:16	Std
7.5		36918	-	7.5	ISM02.4	PPB	2/5/2019 10:18	Std
10		48555	-	10	ISM02.4	PPB	2/5/2019 10:26	Std
ICV	ICV	18860	3.8214 -		ISM02.4	PPB	2/5/2019 10:31	SMPL
ICB	ICB	231	-0.0189 -		ISM02.4	PPB	2/5/2019 10:33	SMPL
CCV075	CCV075	24229	4.9282 -		ISM02.4	PPB	2/5/2019 10:35	SMPL
CCB075	CCB075	6	-0.0653 -		ISM02.4	PPB	2/5/2019 10:38	SMPL
PB116843BL	PBS843	391	0.0141 -		ISM02.4	PPB	2/5/2019 10:40	SMPL
K1311-01	MCB585	1318	0.2052 -		ISM02.4	PPB	2/5/2019 10:42	SMPL
K1311-02	MCB585D	1313	0.2041 -		ISM02.4	PPB	2/5/2019 10:45	SMPL
K1311-03	MCB585S	16397	3.3137 -		ISM02.4	PPB	2/5/2019 10:47	SMPL
K1311-04	MCB586	907	0.1204 -		ISM02.4	PPB	2/5/2019 10:49	SMPL
K1311-05	MCB599	1345	0.2107 -		ISM02.4	PPB	2/5/2019 10:51	SMPL
K1311-06	MCB5A0	1370	0.2159 -		ISM02.4	PPB	2/5/2019 10:54	SMPL
K1311-07	MCB5A1	359	0.0075 -		ISM02.4	PPB	2/5/2019 10:56	SMPL
CCV076	CCV076	23938	4.8682 -		ISM02.4	PPB	2/5/2019 10:58	SMPL
CCB076	CCB076	-9	-0.0684 -		ISM02.4	PPB	2/5/2019 11:00	SMPL