



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/6/2019

OVENTEMP IN Celsius(°C): 108  
Time IN: 14:40  
In Date: 02/05/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:30  
Out Date: 02/06/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-1  
Thermometer ID: %SOLIDS-OVEN

QC:LB100655

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K1339-01 | MGAFY6           | 1     | 1.12              | 9.58                          | 7.73                              | 78.1    |
| K1339-02 | MGAFY7           | 2     | 1.14              | 9.9                           | 8.51                              | 84.1    |
| K1339-03 | MGAFY8           | 3     | 1.12              | 9.82                          | 8.23                              | 81.7    |
| K1339-04 | MGAFY9           | 4     | 1.16              | 9.81                          | 7.96                              | 78.6    |
| K1339-05 | MGAFZ0           | 5     | 1.18              | 9.88                          | 7.2                               | 69.2    |
| K1339-06 | MGAFZ0D          | 6     | 1.18              | 9.88                          | 7.2                               | 69.2    |
| K1339-07 | MGAFZ0S          | 7     | 1.18              | 9.88                          | 7.2                               | 69.2    |
| K1339-08 | MGAFZ1           | 8     | 1.14              | 9.96                          | 7.89                              | 76.5    |
| K1339-09 | MGAFZ2           | 9     | 1.16              | 9.85                          | 8.00                              | 78.7    |
| K1339-10 | MGAFZ3           | 10    | 1.11              | 9.5                           | 8.5                               | 88.1    |
| K1339-11 | MGAFZ4           | 11    | 1.19              | 9.97                          | 8.82                              | 86.9    |
| K1339-12 | MGAFZ5           | 12    | 1.18              | 9.82                          | 7.57                              | 74.0    |
| K1339-13 | MGAFZ6           | 13    | 1.13              | 9.59                          | 8.07                              | 82.0    |
| K1339-14 | MGAFZ7           | 14    | 1.13              | 9.89                          | 8.22                              | 80.9    |
| K1339-15 | MGAFZ8           | 15    | 1.17              | 9.78                          | 8.05                              | 79.9    |
| K1339-16 | MGAFZ9           | 16    | 1.16              | 9.93                          | 8.68                              | 85.7    |
| K1339-17 | MGAG00           | 17    | 1.1               | 9.59                          | 7.73                              | 78.1    |
| K1339-18 | MGAG01           | 18    | 1.14              | 9.87                          | 8.21                              | 81.0    |
| K1339-19 | MGAG02           | 19    | 1.15              | 9.71                          | 8.79                              | 89.3    |
| K1339-20 | MGAG03           | 20    | 1.1               | 9.83                          | 8.5                               | 84.8    |
| K1339-21 | MGAG04           | 21    | 1.15              | 9.95                          | 7.64                              | 73.8    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$