



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB100665

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP71440
COD calibration std. 100 ppm	WP71438
COD calibration std. 10 ppm	WP71436
COD calibration std. 150 ppm	WP71439
COD calibration std. 50 ppm	WP71437
COD calibration std. 0 ppm	WP71435
COD CCV std, 50ppm	WP71441
COD Digestion Vials Low Level 0-150Mg/L	W2444

Temp In (C): 150	Date In: 02/05/2019	Time In: 10:30
Temp Out (C): 150	Date Out: 02/05/2019	Time Out: 12:28

Intercept: -0.1952

Slope: 1.0031

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	02/05/2019	13:30
2	CAL2	10	1	Water	10.000	02/05/2019	13:30
3	CAL3	50	1	Water	49.000	02/05/2019	13:31
4	CAL4	100	1	Water	101.000	02/05/2019	13:31
5	CAL5	150	1	Water	150.000	02/05/2019	13:32



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB100665

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.040	02/05/2019	13:32
2	ICB				1	Water	0.000	0.195	02/05/2019	13:33
3	CCV1	50			1	Water	50.000	50.040	02/05/2019	13:33
4	CCB1				1	Water	0.000	0.195	02/05/2019	13:34
5	LB100665BL				1	Water	0.000	0.195	02/05/2019	13:34
6	LB100665BS	50			1	Water	51.000	51.037	02/05/2019	13:35
7	K1359-03				100	Water	48.000	48.046	02/05/2019	13:35
8	K1377-06				50	Water	88.000	87.923	02/05/2019	13:36
9	K1377-06DUP				50	Water	88.000	87.923	02/05/2019	13:36
10	K1377-06MS	50			50	Water	89.000	88.920	02/05/2019	13:37
11	K1377-06MSD	50			50	Water	89.000	88.920	02/05/2019	13:37
12	CCV2	50			1	Water	51.000	51.037	02/05/2019	13:38
13	CCB2				1	Water	0.000	0.195	02/05/2019	13:38

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CN2419

WorkList ID : 121886

Date : 2/4/2019 9:16:50 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/12/2019	Solid	K1284-01	A Cyanide	Cool 4 deg C	TULL02	L52	FK-BPM-03-004-ABCD	01/30/2019	9012B
02/06/2019	Solid	K1312-01	J Cyanide	Cool 4 deg C	PSEG01	K31	SOIL-PILE	02/01/2019	9012B
02/06/2019	Solid	K1360-01	K Cyanide	Cool 4 deg C	PSEG01	K23	RBR-200053	02/01/2019	9012B
02/06/2019	Solid	K1360-03	K Cyanide	Cool 4 deg C	PSEG01	K23	RBR-200034	02/01/2019	9012B
02/15/2019	Solid	K1361-01	B Cyanide	Cool 4 deg C	TULL02	L62	FK-BPM-03-006-ABCD	02/01/2019	9012B
02/15/2019	Solid	K1362-01	B Cyanide	Cool 4 deg C	TULL02	L62	FK-BPM-03-005-ABCD	02/01/2019	9012B

Date/Time 02/04/19 10:15 AM
 Received by: KS
 Relinquished by: JP

Date/Time 02/04/19 10:50 AM
 Received by: JP
 Relinquished by: KS