

LB100666


7470A

INSTRUMENT ID : CU1

Linear

 μ Abs. :

39393



A= 0.0000e+000

B= 2.5600e-004

C= -1.3091e-002

Rho= 0.9998832

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	153	0.000	153				
0.2	0.200	0.191	-0.009	799	0.0 %	799				
2.5	2.500	2.502	0.002	9825	0.0 %	9825				
5.0	5.000	5.022	0.022	19667	0.0 %	19667				
7.5	7.500	7.387	-0.113	28907	0.0 %	28907				
10.0	10.000	10.072	0.072	39393	0.0 %	39393				

LB100666 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Feb 2019 13:38:29

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	153	-	0.0000	7470A	PPB	05 Feb 2019 13:39:57	S	Std
0.2 - 1	50-2	799	-	0.2000	7470A	PPB	05 Feb 2019 13:42:05	S	Std
2.5 - 1	52-5	9825	-	2.5000	7470A	PPB	05 Feb 2019 13:44:13	S	Std
5.0 - 1	55-0	19667	-	5.0000	7470A	PPB	05 Feb 2019 13:46:21	S	Std
7.5 - 1	57-5	28907	-	7.5000	7470A	PPB	05 Feb 2019 13:48:28	S	Std
10.0 - 1	510-0	39393	-	10.0000	7470A	PPB	05 Feb 2019 13:50:35	S	Std
ICV95 - 1	ICV95	14739	3.7601	-	7470A	PPB	05 Feb 2019 13:53:36	U	SMPL
ICB95 - 1	ICB95	1	-0.0128	-	7470A	PPB	05 Feb 2019 13:55:41	U	SMPL
CCV77 - 1	CCV77	19646	5.0163	-	7470A	PPB	05 Feb 2019 13:57:47	U	SMPL
CCB77 - 1	CCB77	5	-0.0118	-	7470A	PPB	05 Feb 2019 13:59:53	U	SMPL
CRA - 1	CRA	789	0.1889	-	7470A	PPB	05 Feb 2019 14:01:59	U	SMPL
HighStd - 1	HighStd	40330	10.3114	-	7470A	PPB	05 Feb 2019 14:04:05	U	SMPL
ChkStd - 1	ChkStd	27801	7.1040	-	7470A	PPB	05 Feb 2019 14:06:11	U	SMPL
PB116826BL - 1	PBW	33	-0.0046	-	7470A	PPB	05 Feb 2019 14:08:19	U	SMPL
PB116826BS - 1	LCSW	14917	3.8057	-	7470A	PPB	05 Feb 2019 14:10:26	U	SMPL
K1359-03 - 1	COMP-3	170	0.0304	-	7470A	PPB	05 Feb 2019 14:12:33	U	SMPL
K1359-03DUP - 1	COMP-3DUP	166	0.0294	-	7470A	PPB	05 Feb 2019 14:14:40	U	SMPL
K1359-03MS - 1	COMP-3MS	16359	4.1748	-	7470A	PPB	05 Feb 2019 14:16:48	U	SMPL
K1359-03MSD - 1	COMP-3MSD	16142	4.1193	-	7470A	PPB	05 Feb 2019 14:18:55	U	SMPL
K1363-02 - 1	INFLUENT	417	0.0937	-	7470A	PPB	05 Feb 2019 14:21:02	U	SMPL
CCV78 - 1	CCV78	19551	4.9920	-	7470A	PPB	05 Feb 2019 14:23:08	U	SMPL
CCB78 - 1	CCB78	15	-0.0093	-	7470A	PPB	05 Feb 2019 14:25:14	U	SMPL
PB116827BL - 1	PBW	114	0.0161	-	7470A	PPB	05 Feb 2019 14:27:20	U	SMPL
PB116827BS - 1	LCSW	15347	3.9158	-	7470A	PPB	05 Feb 2019 14:29:26	U	SMPL
K1296-11 - 1	BP-TT-FB01-20190129	58	0.0018	-	7470A	PPB	05 Feb 2019 14:31:32	U	SMPL
K1296-12 - 1	BP-TT-MW202S-20190129	133	0.0210	-	7470A	PPB	05 Feb 2019 14:33:39	U	SMPL
K1296-13 - 1	BP-TT-RB01-20190129	95	0.0112	-	7470A	PPB	05 Feb 2019 14:35:46	U	SMPL
K1296-14 - 1	BP-TT-MW202I-20190129	109	0.0148	-	7470A	PPB	05 Feb 2019 14:37:53	U	SMPL
K1296-15 - 1	BP-TT-DUP01-20190129	104	0.0135	-	7470A	PPB	05 Feb 2019 14:40:01	U	SMPL
K1296-16 - 1	BP-TT-MW203I-20190130	150	0.0253	-	7470A	PPB	05 Feb 2019 14:42:09	U	SMPL
K1296-16DUP - 1	BP-TT-MW203I-20190130DUP	150	0.0253	-	7470A	PPB	05 Feb 2019 14:44:17	U	SMPL
K1296-16MS - 1	BP-TT-MW203I-20190130MS	15191	3.8758	-	7470A	PPB	05 Feb 2019 14:46:23	U	SMPL
CCV79 - 1	CCV79	19606	5.0061	-	7470A	PPB	05 Feb 2019 14:48:29	U	SMPL
CCB79 - 1	CCB79	-24	-0.0192	-	7470A	PPB	05 Feb 2019 14:50:36	U	SMPL
K1296-16MSD - 1	BP-TT-MW203I-20190130MSD	15050	3.8397	-	7470A	PPB	05 Feb 2019 14:52:42	U	SMPL
PB116828BL - 1	PBW	-22	-0.0187	-	7470A	PPB	05 Feb 2019 14:54:49	U	SMPL
PB116828BS - 1	LCSW	15922	4.0630	-	7470A	PPB	05 Feb 2019 15:01:24	U	SMPL
K1359-01 - 1	COMP-1	43	-0.0021	-	7470A	PPB	05 Feb 2019 15:03:29	U	SMPL
K1359-01DUP - 1	COMP-1DUP	26	-0.0064	-	7470A	PPB	05 Feb 2019 15:05:37	U	SMPL
K1359-01MS - 1	COMP-1MS	15411	3.9321	-	7470A	PPB	05 Feb 2019 15:07:44	U	SMPL
K1359-01MSD - 1	COMP-1MSD	15430	3.9370	-	7470A	PPB	05 Feb 2019 15:09:51	U	SMPL
K1359-02 - 1	COMP-2	-41	-0.0236	-	7470A	PPB	05 Feb 2019 15:11:59	U	SMPL
K1359-04 - 1	COMP-3	21	-0.0077	-	7470A	PPB	05 Feb 2019 15:14:07	U	SMPL
PB116829BL - 1	PBW	23	-0.0072	-	7470A	PPB	05 Feb 2019 15:16:13	U	SMPL
CCV80 - 1	CCV80	19840	5.0660	-	7470A	PPB	05 Feb 2019 15:18:20	U	SMPL
CCB80 - 1	CCB80	-14	-0.0167	-	7470A	PPB	05 Feb 2019 15:20:27	U	SMPL
PB116829BS - 1	LCSW	14442	3.6841	-	7470A	PPB	05 Feb 2019 15:22:33	U	SMPL
K1312-02 - 1	SOIL-PILE	-20	-0.0182	-	7470A	PPB	05 Feb 2019 15:24:39	U	SMPL
K1312-02DUP - 1	SOIL-PILEDUP	2	-0.0126	-	7470A	PPB	05 Feb 2019 15:26:46	U	SMPL
K1312-02MS - 1	SOIL-PILEMS	14506	3.7005	-	7470A	PPB	05 Feb 2019 15:28:54	U	SMPL
K1312-02MSD - 1	SOIL-PILEMSD	14768	3.7675	-	7470A	PPB	05 Feb 2019 15:31:00	U	SMPL
K1334-03 - 1	50151-TOTE	767	0.1833	-	7470A	PPB	05 Feb 2019 15:33:08	U	SMPL
K1360-02 - 1	RBR-200053	48	-0.0008	-	7470A	PPB	05 Feb 2019 15:35:15	U	SMPL
K1360-04 - 1	RBR-200034	44	-0.0018	-	7470A	PPB	05 Feb 2019 15:37:22	U	SMPL
PB116828TB - 1	PB116781TB	50	-0.0003	-	7470A	PPB	05 Feb 2019 15:39:30	U	SMPL
PB116829TB - 1	PB116779TB	34	-0.0044	-	7470A	PPB	05 Feb 2019 15:41:37	U	SMPL
CCV81 - 1	CCV81	19510	4.9815	-	7470A	PPB	05 Feb 2019 15:43:43	U	SMPL
CCB81 - 1	CCB81	16	-0.0090	-	7470A	PPB	05 Feb 2019 15:45:52	U	SMPL
K1359-03LX5 - 1		33	-0.0046	-	7470A	PPB	05 Feb 2019 15:47:58	U	SMPL
K1359-03A - 1		16109	4.1108	-	7470A	PPB	05 Feb 2019 15:50:04	U	SMPL
K1296-16LX5 - 1		67	0.0041	-	7470A	PPB	05 Feb 2019 15:52:10	U	SMPL
K1296-16A - 1		15311	3.9065	-	7470A	PPB	05 Feb 2019 15:54:17	U	SMPL
K1359-01LX5 - 1		22	-0.0075	-	7470A	PPB	05 Feb 2019 15:56:24	U	SMPL
K1359-01A - 1		14607	3.7263	-	7470A	PPB	05 Feb 2019 15:58:31	U	SMPL
K1312-02LX5 - 1		35	-0.0041	-	7470A	PPB	05 Feb 2019 16:00:38	U	SMPL
K1312-02A - 1		14975	3.8205	-	7470A	PPB	05 Feb 2019 16:02:46	U	SMPL
CCV82 - 1	CCV82	19074	4.8699	-	7470A	PPB	05 Feb 2019 16:04:54	U	SMPL
CCB82 - 1	CCB82	-39	-0.0231	-	7470A	PPB	05 Feb 2019 16:07:01	U	SMPL