



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100675

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	02/05/2019	17:00
2	CAL2	1	Water	NA	NA	21.3	7.00	02/05/2019	17:01
3	CAL3	1	Water	NA	NA	21.3	10.04	02/05/2019	17:03
4	ICV	1	Water	NA	NA	21.2	7.00	02/05/2019	17:05
5	CCV1	1	Water	NA	NA	21.2	2.02	02/05/2019	17:06
6	K1338-02	1	Solid	20.02	20	21.2	7.25	02/05/2019	17:10
7	K1338-02DUP	1	Solid	20.03	20	21.3	7.26	02/05/2019	17:11
8	K1338-04	1	Solid	20.04	20	21.1	8.46	02/05/2019	17:12
9	K1338-06	1	Solid	20.02	20	21.0	7.84	02/05/2019	17:13
10	K1356-02	1	Solid	20.04	20	21.9	8.43	02/05/2019	17:15
11	K1356-04	1	Solid	20.02	20	21.7	7.81	02/05/2019	17:16
12	K1358-02	1	Solid	20.03	20	21.3	7.17	02/05/2019	17:19
13	K1379-02	1	Solid	20.02	20	21.6	7.45	02/05/2019	17:20
14	CCV2	1	Water	NA	NA	21.2	12.01	02/05/2019	17:21

2B 100675

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity k1379

WorkList ID : 121972

Date : 02-05-2019 15:17:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2019	Solid	K1338-02	Corrosivity	Cool 4 deg C	PSEG05	K13	AU-06-020519-A	02/05/2019	9045D
02/05/2019	Solid	K1338-04	Corrosivity	Cool 4 deg C	PSEG05	K13	AU-06-020519-B	02/05/2019	9045D
02/05/2019	Solid	K1338-06	Corrosivity	Cool 4 deg C	PSEG05	K13	AU-06-020519-C	02/05/2019	9045D
02/05/2019	Solid	K1356-02	Corrosivity	Cool 4 deg C	PSEG05	K13	SU-01-020519-A	02/05/2019	9045D
02/05/2019	Solid	K1356-04	Corrosivity	Cool 4 deg C	PSEG05	K13	SU-01-020519-B	02/05/2019	9045D
02/05/2019	Solid	K1358-02	Corrosivity	Cool 4 deg C	PSEG05	K31	TR-05-020519	02/05/2019	9045D
02/05/2019	Solid	K1379-02	Corrosivity	Cool 4 deg C	PSEG01	K31	HR-6A	02/05/2019	9045D

Date/Time 02-05-19 16:20 PM
 Received by: Jo
 Relinquished by: CG

Date/Time 02-05-19 6:25 PM
 Received by: Jo
 Relinquished by: Jo