

LB100677


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55902



A= 0.0000e+000

B= 1.7868e-004

C= -2.9685e-002

Rho= 0.9999242

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.042	-0.042	-68	0.000	-68				
0.2	0.200	0.190	-0.010	1227	0.0 %	1227				
2.5	2.500	2.575	0.075	14577	0.0 %	14577				
5.0	5.000	4.970	-0.030	27981	0.0 %	27981				
7.5	7.500	7.548	0.048	42410	0.0 %	42410				
10.0	10.000	9.959	-0.041	55902	0.0 %	55902				

LB100677 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Feb 2019 16:57:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-68	-	0.0000	7471B	PPB	05 Feb 2019 17:18:26	S	Std
0.2 - 1	50.2	1227	-	0.2000	7471B	PPB	05 Feb 2019 17:20:43	S	Std
2.5 - 1	52.5	14577	-	2.5000	7471B	PPB	05 Feb 2019 17:23:00	S	Std
5.0 - 1	55.0	27981	-	5.0000	7471B	PPB	05 Feb 2019 17:25:17	S	Std
7.5 - 1	57.5	42410	-	7.5000	7471B	PPB	05 Feb 2019 17:27:34	S	Std
10.0 - 1	510.0	55902	-	10.0000	7471B	PPB	05 Feb 2019 17:29:51	S	Std
ICV96 - 1	ICV96	22133	3.9251	-7471B	PPB	PPB	05 Feb 2019 17:34:01	U	SMPL
ICB96 - 1	ICB96	-61	-0.0406	-7471B	PPB	PPB	05 Feb 2019 17:36:16	U	SMPL
CCV83 - 1	CCV83	28377	5.0408	-7471B	PPB	PPB	05 Feb 2019 17:38:32	U	SMPL
CCB83 - 1	CCB83	-53	-0.0392	-7471B	PPB	PPB	05 Feb 2019 17:40:49	U	SMPL
CRA - 1	CRA	1230	0.1901	-7471B	PPB	PPB	05 Feb 2019 17:46:55	U	SMPL
HighStd - 1	HighStd	56054	9.9862	-7471B	PPB	PPB	05 Feb 2019 17:49:10	U	SMPL
ChkStd - 1	ChkStd	40788	7.2584	-7471B	PPB	PPB	05 Feb 2019 17:51:25	U	SMPL
PB116858BL - 1	PBS	-37	-0.0363	-7471B	PPB	PPB	05 Feb 2019 17:53:43	U	SMPL
PB116858BS - 1	LCCS	20996	3.7219	-7471B	PPB	PPB	05 Feb 2019 17:59:56	U	SMPL
K1284-01 - 1	FK-BPM-03-004-ABCD	1623	0.2603	-7471B	PPB	PPB	05 Feb 2019 18:02:13	U	SMPL
K1284-01DUP - 1	FK-BPM-03-004-ABCD DUP	1651	0.2653	-7471B	PPB	PPB	05 Feb 2019 18:04:31	U	SMPL
K1284-01MS - 1	FK-BPM-03-004-ABCDMS	23428	4.1565	-7471B	PPB	PPB	05 Feb 2019 18:10:34	U	SMPL
K1284-01MSD - 1	FK-BPM-03-004-ABCDMSD	23306	4.1347	-7471B	PPB	PPB	05 Feb 2019 18:12:51	U	SMPL
K1312-01 - 1	SOIL-PILE	574	0.0729	-7471B	PPB	PPB	05 Feb 2019 18:15:07	U	SMPL
CCV84 - 1	CCV84	28084	4.9884	-7471B	PPB	PPB	05 Feb 2019 18:17:23	U	SMPL
CCB84 - 1	CCB84	-40	-0.0368	-7471B	PPB	PPB	05 Feb 2019 18:19:38	U	SMPL
K1341-01 - 1	HR-05-020419-A	4264	0.7322	-7471B	PPB	PPB	05 Feb 2019 18:21:55	U	SMPL
K1341-03 - 1	HR-05-020419-B	3346	0.5682	-7471B	PPB	PPB	05 Feb 2019 18:24:10	U	SMPL
K1341-05 - 1	HR-05-020419-C	5691	0.9872	-7471B	PPB	PPB	05 Feb 2019 18:26:26	U	SMPL
K1346-01 - 1	CL-02-020419-A	1851	0.3011	-7471B	PPB	PPB	05 Feb 2019 18:28:41	U	SMPL
K1346-03 - 1	CL-02-020419-B	3041	0.5137	-7471B	PPB	PPB	05 Feb 2019 18:30:58	U	SMPL
K1346-05 - 1	CL-02-020419-C	5602	0.9713	-7471B	PPB	PPB	05 Feb 2019 18:33:14	U	SMPL
K1350-01 - 1	EO-03-020419-A	2568	0.4292	-7471B	PPB	PPB	05 Feb 2019 18:35:31	U	SMPL
K1350-03 - 1	EO-03-020419-B	1233	0.1906	-7471B	PPB	PPB	05 Feb 2019 18:37:48	U	SMPL
K1360-01 - 1	RBR-200053	1040	0.1561	-7471B	PPB	PPB	05 Feb 2019 18:40:05	U	SMPL
K1360-03 - 1	RBR-200034	3221	0.5459	-7471B	PPB	PPB	05 Feb 2019 18:42:21	U	SMPL
CCV85 - 1	CCV85	28170	5.0038	-7471B	PPB	PPB	05 Feb 2019 18:44:37	U	SMPL
CCB85 - 1	CCB85	-90	-0.0458	-7471B	PPB	PPB	05 Feb 2019 18:46:52	U	SMPL
K1361-01 - 1	FK-BPM-03-006-ABCD	1627	0.2610	-7471B	PPB	PPB	05 Feb 2019 18:49:09	U	SMPL
K1362-01 - 1	FK-BPM-03-005-ABCD	1815	0.2946	-7471B	PPB	PPB	05 Feb 2019 18:51:25	U	SMPL
K1284-01LX5 - 1		306	0.0250	-7471B	PPB	PPB	05 Feb 2019 18:53:40	U	SMPL
K1284-01A - 1		22428	3.9778	-7471B	PPB	PPB	05 Feb 2019 18:55:56	U	SMPL
CCV86 - 1	CCV86	27781	4.9343	-7471B	PPB	PPB	05 Feb 2019 18:58:12	U	SMPL
CCB86 - 1	CCB86	-94	-0.0465	-7471B	PPB	PPB	05 Feb 2019 19:00:29	U	SMPL