



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100681

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.2	4.01	02/06/2019	09:30
2	CAL2	1	Water	NA	NA	22.1	7.01	02/06/2019	09:31
3	CAL3	1	Water	NA	NA	22.2	10.04	02/06/2019	09:32
4	ICV	1	Water	NA	NA	22.1	6.99	02/06/2019	09:34
5	CCV1	1	Water	NA	NA	22.1	2.01	02/06/2019	09:35
6	K1383-04	1	Water	NA	NA	19.5	5.70	02/06/2019	09:36
7	K1383-04DUP	1	Water	NA	NA	19.6	5.71	02/06/2019	09:37
8	CCV2	1	Water	NA	NA	22.1	12.01	02/06/2019	09:38

2B100681

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1338

WorkList ID : 121995

Date : 02-06-2019 09:17:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2019	Water	K1383-04-1	pH	Cool 4 deg C	WEST10	L41	FB-02-05-19	02/05/2019	9040C

Date/Time 02-06-19 09:20 AM

Received by: J

Relinquished by: PC

Date/Time 020619 10:30 AM

Received by: PC

Relinquished by: SQ