



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB100682

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.2	4.01	02/06/2019	09:30
2	CAL2	1	Water	NA	NA	22.1	7.01	02/06/2019	09:31
3	CAL3	1	Water	NA	NA	22.2	10.04	02/06/2019	09:32
4	ICV	1	Water	NA	NA	22.1	6.99	02/06/2019	09:34
5	CCV1	1	Water	NA	NA	22.2	2.02	02/06/2019	09:40
6	K1383-01	1	Solid	20.03	20	22.3	7.83	02/06/2019	09:41
7	K1383-01DUP	1	Solid	20.04	20	22.4	7.84	02/06/2019	09:42
8	K1383-02	1	Solid	20.02	20	21.7	7.70	02/06/2019	09:45
9	K1383-03	1	Solid	20.04	20	20.7	7.53	02/06/2019	09:46
10	CCV2	1	Water	NA	NA	22.1	12.01	02/06/2019	09:47

LB 100682

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K1338

WorkList ID : 121996

Date : 02-06-2019 09:17:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/05/2019	Solid	K1383-01 ~ [	pH	Cool 4 deg C	WEST10	L41	A0C5B-SS1-0.0-0.5	02/05/2019	9045D
02/05/2019	Solid	K1383-02 ~ [	pH	Cool 4 deg C	WEST10	L41	A0C5B-SS1-4.5-5.0	02/05/2019	9045D
02/05/2019	Solid	K1383-03 ~ [	pH	Cool 4 deg C	WEST10	L41	A0C5B-SS2-0.0-0.5	02/05/2019	9045D

Date/Time 02.06.19 09.20 AM

Received by: Jb

Relinquished by: PC

Date/Time

Received by: PC

Relinquished by: Jb