


LB100684
INSTRUMENT ID: CV1

7470A

Linear



A= 0.0000e+000

B= 2.5259e-004

C= -4.5822e-002

Rho= 0.9999422

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	51	0.000	51				
0.2	0.200	0.173	-0.027	866	0.0 %	866				
2.5	2.500	2.538	0.038	10228	0.0 %	10228				
5.0	5.000	5.052	0.052	20184	0.0 %	20184				
7.5	7.500	7.524	0.024	29967	0.0 %	29967				
10.0	10.000	9.946	-0.054	39558	0.0 %	39558				

LB100684 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Feb 2019 10:56:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	51	-	0.00007470A	PPB		06 Feb 2019 11:02:28	S	Std
0.2 - 1	50-2	866	-	0.20007470A	PPB		06 Feb 2019 11:04:36	S	Std
2.5 - 1	52-5	10228	-	2.50007470A	PPB		06 Feb 2019 11:06:44	S	Std
5.0 - 1	55-0	20184	-	5.00007470A	PPB		06 Feb 2019 11:08:52	S	Std
7.5 - 1	57-5	29967	-	7.50007470A	PPB		06 Feb 2019 11:10:59	S	Std
10.0 - 1	510-0	39558	-	10.00007470A	PPB		06 Feb 2019 11:13:05	S	Std
ICV97 - 1	ICV97	15595	3.8934	-7470A	PPB		06 Feb 2019 11:18:25	U	SMPL
ICB97 - 1	ICB97	9	-0.0435	-7470A	PPB		06 Feb 2019 11:20:31	U	SMPL
CCV87 - 1	CCV87	20036	5.0151	-7470A	PPB		06 Feb 2019 11:22:36	U	SMPL
CCB87 - 1	CCB87	-6	-0.0473	-7470A	PPB		06 Feb 2019 11:24:42	U	SMPL
CRA - 1	CRA	895	0.1802	-7470A	PPB		06 Feb 2019 11:26:48	U	SMPL
HighStd - 1	HighStd	39992	10.0558	-7470A	PPB		06 Feb 2019 11:28:54	U	SMPL
ChkStd - 1	ChkStd	28384	7.1238	-7470A	PPB		06 Feb 2019 11:31:01	U	SMPL
PB116878BL - 1	PBW	0	-0.0458	-7470A	PPB		06 Feb 2019 11:33:08	U	SMPL
PB116878BS - 1	LCSW	15670	3.9123	-7470A	PPB		06 Feb 2019 11:35:15	U	SMPL
K1341-02 - 1	HR-05-020419-A	9	-0.0435	-7470A	PPB		06 Feb 2019 11:37:22	U	SMPL
K1341-02DUP - 1	HR-05-020419-ADUP	30	-0.0382	-7470A	PPB		06 Feb 2019 11:39:29	U	SMPL
K1341-02MS - 1	HR-05-020419-AMS	14480	3.6117	-7470A	PPB		06 Feb 2019 11:41:37	U	SMPL
K1341-02MSD - 1	HR-05-020419-AMSD	14605	3.6433	-7470A	PPB		06 Feb 2019 11:43:46	U	SMPL
K1341-04 - 1	HR-05-020419-B	74	-0.0271	-7470A	PPB		06 Feb 2019 11:45:52	U	SMPL
CCV88 - 1	CCV88	20183	5.0522	-7470A	PPB		06 Feb 2019 11:47:58	U	SMPL
CCB88 - 1	CCB88	-6	-0.0473	-7470A	PPB		06 Feb 2019 11:50:04	U	SMPL
K1341-06 - 1	HR-05-020419-C	149	-0.0082	-7470A	PPB		06 Feb 2019 11:52:10	U	SMPL
K1346-02 - 1	CL-02-020419-A	98	-0.0211	-7470A	PPB		06 Feb 2019 11:54:16	U	SMPL
K1346-04 - 1	CL-02-020419-B	42	-0.0352	-7470A	PPB		06 Feb 2019 11:56:22	U	SMPL
K1346-06 - 1	CL-02-020419-C	50	-0.0332	-7470A	PPB		06 Feb 2019 11:58:28	U	SMPL
K1350-02 - 1	EO-03-020419-A	44	-0.0347	-7470A	PPB		06 Feb 2019 12:00:35	U	SMPL
K1350-04 - 1	EO-03-020419-B	55	-0.0319	-7470A	PPB		06 Feb 2019 12:02:42	U	SMPL
B116878TB - 1	B116823TB	139	-0.0107	-7470A	PPB		06 Feb 2019 12:04:50	U	SMPL
K1341-02LX5 - 1		82	-0.0251	-7470A	PPB		06 Feb 2019 12:06:57	U	SMPL
K1341-02A - 1		15767	3.9368	-7470A	PPB		06 Feb 2019 12:09:06	U	SMPL
CCV89 - 1	CCV89	20276	5.0757	-7470A	PPB		06 Feb 2019 12:11:12	U	SMPL
CCB89 - 1	CCB89	7	-0.0441	-7470A	PPB		06 Feb 2019 12:13:18	U	SMPL