


LB100687

INSTRUMENT ID: CV1

7471B

μ Abs.:

55222



Linear

A= 0.0000e+000

B= 1.8012e-004

C= -7.1210e-002

Rho= 0.9997734

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.067	-0.067	23	0.000	23				
0.2	0.200	0.184	-0.016	1419	0.0 %	1419				
2.5	2.500	2.530	0.030	14443	0.0 %	14443				
5.0	5.000	5.077	0.077	28579	0.0 %	28579				
7.5	7.500	7.600	0.100	42589	0.0 %	42589				
10.0	10.000	9.876	-0.124	55222	0.0 %	55222				

LB100687 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Feb 2019 13:30:36

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50-0	23	-	0.00007471B	PPB	06 Feb 2019 13:43:34	S	Std
0.2 - 1	50-2	1419	-	0.20007471B	PPB	06 Feb 2019 13:45:51	S	Std
2.5 - 1	52-5	14443	-	2.50007471B	PPB	06 Feb 2019 13:48:09	S	Std
5.0 - 1	55-0	28579	-	5.00007471B	PPB	06 Feb 2019 13:50:26	S	Std
7.5 - 1	57-5	42589	-	7.50007471B	PPB	06 Feb 2019 13:52:42	S	Std
10.0 - 1	510-0	55222	-	10.00007471B	PPB	06 Feb 2019 13:58:27	S	Std
ICV98 - 1	ICV98	22343	3.9533	-7471B	PPB	06 Feb 2019 14:03:06	U	SMPL
ICB98 - 1	ICB98	-19	-0.0746	-7471B	PPB	06 Feb 2019 14:05:22	U	SMPL
CCV90 - 1	CCV90	28761	5.1094	-7471B	PPB	06 Feb 2019 14:07:38	U	SMPL
CCB90 - 1	CCB90	-40	-0.0784	-7471B	PPB	06 Feb 2019 14:09:53	U	SMPL
CRA - 1	CRA	1409	0.1826	-7471B	PPB	06 Feb 2019 14:12:09	U	SMPL
HighStd - 1	HighStd	56432	10.0936	-7471B	PPB	06 Feb 2019 14:14:24	U	SMPL
ChkStd - 1	ChkStd	40092	7.1504	-7471B	PPB	06 Feb 2019 14:16:40	U	SMPL
PB116869BL - 1	PBS	192	-0.0366	-7471B	PPB	06 Feb 2019 14:18:57	U	SMPL
PB116869BS - 1	LCSS	22435	3.9699	-7471B	PPB	06 Feb 2019 14:24:53	U	SMPL
K1338-01 - 1	AU-06-020519-A	2422	0.3651	-7471B	PPB	06 Feb 2019 14:27:09	U	SMPL
K1338-03 - 1	AU-06-020519-B	2232	0.3308	-7471B	PPB	06 Feb 2019 14:29:26	U	SMPL
K1338-05 - 1	AU-06-020519-C	1693	0.2337	-7471B	PPB	06 Feb 2019 14:31:44	U	SMPL
K1356-01 - 1	SU-01-020519-A	1109	0.1285	-7471B	PPB	06 Feb 2019 14:34:01	U	SMPL
K1356-03 - 1	SU-01-020519-B	3875	0.6268	-7471B	PPB	06 Feb 2019 14:36:17	U	SMPL
CCV91 - 1	CCV91	28971	5.1472	-7471B	PPB	06 Feb 2019 14:38:32	U	SMPL
CCB91 - 1	CCB91	19	-0.0678	-7471B	PPB	06 Feb 2019 14:40:48	U	SMPL
K1358-01 - 1	TR-05-020519	1491	0.1974	-7471B	PPB	06 Feb 2019 14:43:04	U	SMPL
K1379-01 - 1	HR-6A	109178	19.5945	-7471B	PPB	06 Feb 2019 14:45:20	U	SMPL
K1379-01DUP - 1	HR-6ADUP	110939	19.9117	-7471B	PPB	06 Feb 2019 14:47:36	U	SMPL
K1379-01MS - 1	HR-6AMS	119341	21.4251	-7471B	PPB	06 Feb 2019 14:49:54	U	SMPL
K1379-01MSD - 1	HR-6AMSD	119572	21.4667	-7471B	PPB	06 Feb 2019 14:57:37	U	SMPL
K1379-01LX5 - 1		21875	3.8690	-7471B	PPB	06 Feb 2019 14:59:53	U	SMPL
K1379-01A - 1		129660	23.2838	-7471B	PPB	06 Feb 2019 15:02:12	U	SMPL
K1379-01DLX2 - 1		52518	9.3886	-7471B	PPB	06 Feb 2019 15:04:29	U	SMPL
K1379-01DUPDLX2 - 1		52263	9.3427	-7471B	PPB	06 Feb 2019 15:06:49	U	SMPL
CCV92 - 1	CCV92	28723	5.1025	-7471B	PPB	06 Feb 2019 15:09:06	U	SMPL
CCB92 - 1	CCB92	-26	-0.0759	-7471B	PPB	06 Feb 2019 15:11:24	U	SMPL
K1379-01MSDLX4 - 1		29309	5.2081	-7471B	PPB	06 Feb 2019 15:37:42	U	SMPL
K1379-01MSDDLX4 - 1		28944	5.1423	-7471B	PPB	06 Feb 2019 15:39:58	U	SMPL
K1379-01LX5DLX2 - 1		10637	1.8448	-7471B	PPB	06 Feb 2019 15:42:14	U	SMPL
K1379-01ADLX4 - 1		30967	5.5067	-7471B	PPB	06 Feb 2019 15:44:31	U	SMPL
CCV93 - 1	CCV93	28958	5.1448	-7471B	PPB	06 Feb 2019 15:46:47	U	SMPL
CCB93 - 1	CCB93	-51	-0.0804	-7471B	PPB	06 Feb 2019 15:49:04	U	SMPL