



Analytical Summary Report

Analysis Method: 1030
Parameter: Ignitability
Run Number: LB100689

Reviewed By: ketankumar
Supervisor Review By: apatel

Seq	LabID	ClientID	DF	matrix	Result Status	Burning Rate	Anal Date	Anal Time
1	K1338-02	AU-06-020519-A	1	Solid	NO	0.00	02/06/2019	09:45
2	K1338-02DUP	AU-06-020519-ADUP	1	Solid	NO	0.00	02/06/2019	09:58
3	K1338-04	AU-06-020519-B	1	Solid	NO	0.00	02/06/2019	10:10
4	K1338-06	AU-06-020519-C	1	Solid	NO	0.00	02/06/2019	10:20
5	K1356-02	SU-01-020519-A	1	Solid	NO	0.00	02/06/2019	10:32
6	K1356-04	SU-01-020519-B	1	Solid	NO	0.00	02/06/2019	10:42
7	K1358-02	TR-05-020519	1	Solid	NO	0.00	02/06/2019	10:50
8	K1379-02	HR-6A	1	Solid	NO	0.00	02/06/2019	11:02

Burning Rate = $\frac{\text{Length (mm)}}{\text{Total Time (sec)}}$

WORKLIST(Hardcopy Internal Chain)

LB100689

WorkList Name : IGNI2619

WorkList ID : 122001

Date : 02-06-2019 09:28:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/08/2019	Solid	K1338-02	Ignitability	Cool 4 deg C	PSEG05	K13	AU-06-020519-A	02/05/2019	1030
02/08/2019	Solid	K1338-04	Ignitability	Cool 4 deg C	PSEG05	K13	AU-06-020519-B	02/05/2019	1030
02/08/2019	Solid	K1338-06	Ignitability	Cool 4 deg C	PSEG05	K13	AU-06-020519-C	02/05/2019	1030
02/08/2019	Solid	K1356-02	Ignitability	Cool 4 deg C	PSEG05	K13	SU-01-020519-A	02/05/2019	1030
02/08/2019	Solid	K1356-04	Ignitability	Cool 4 deg C	PSEG05	K13	SU-01-020519-B	02/05/2019	1030
02/08/2019	Solid	K1358-02	Ignitability	Cool 4 deg C	PSEG05	K31	TR-05-020519	02/05/2019	1030
02/08/2019	Solid	K1379-02	Ignitability	Cool 4 deg C	PSEG01	K31	HR-6A	02/05/2019	1030

Date/Time 02/06/19 9:30
 Received by: KS
 Relinquished by: CP

Date/Time 02/06/19 12:40
 Received by: CP
 Relinquished by: KS