



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB100696

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/06/2019	17:20
2	CAL2	1	Water	NA	NA	22.1	7.01	02/06/2019	17:21
3	CAL3	1	Water	NA	NA	22.1	10.03	02/06/2019	17:22
4	ICV	1	Water	NA	NA	22.1	7.00	02/06/2019	17:24
5	CCV1	1	Water	NA	NA	22.2	2.02	02/06/2019	17:25
6	K1389-21	1	Water	NA	NA	10.2	6.60	02/06/2019	17:26
7	K1389-21DUP	1	Water	NA	NA	10.3	6.61	02/06/2019	17:27
8	K1396-04	1	Water	NA	NA	13.1	6.85	02/06/2019	17:30
9	K1396-05	1	Water	NA	NA	13.4	7.18	02/06/2019	17:31
10	K1396-06	1	Water	NA	NA	13.5	6.95	02/06/2019	17:32
11	K1396-07	1	Water	NA	NA	14.3	4.93	02/06/2019	17:33
12	K1396-08	1	Water	NA	NA	14.5	6.80	02/06/2019	17:35
13	K1396-09	1	Water	NA	NA	14.6	7.43	02/06/2019	17:36
14	K1396-10	1	Water	NA	NA	14.7	7.30	02/06/2019	17:37
15	K1396-11	1	Water	NA	NA	15.4	6.65	02/06/2019	17:39
16	CCV2	1	Water	NA	NA	22.3	12.02	02/06/2019	17:40
17	K1396-12	1	Water	NA	NA	14.8	6.08	02/06/2019	17:41
18	CCV3	1	Water	NA	NA	22.2	2.02	02/06/2019	17:42

13100696

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K1389

WorkList ID : 122021

Date : 02-06-2019 15:09:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2019	Water	K1389-21 -1	pH	Cool 4 deg C	PSEG01	K31	M-7	02/06/2019	9040C
02/06/2019	Water	K1396-04	pH	Cool 4 deg C	PSEG01	K31	23866	02/06/2019	9040C
02/06/2019	Water	K1396-05	pH	Cool 4 deg C	PSEG01	K31	23867	02/06/2019	9040C
02/06/2019	Water	K1396-06	pH	Cool 4 deg C	PSEG01	K31	23890	02/06/2019	9040C
02/06/2019	Water	K1396-07	pH	Cool 4 deg C	PSEG01	K31	23918	02/06/2019	9040C
02/06/2019	Water	K1396-08	pH	Cool 4 deg C	PSEG01	K31	23919	02/06/2019	9040C
02/06/2019	Water	K1396-09	pH	Cool 4 deg C	PSEG01	K31	23936	02/06/2019	9040C
02/06/2019	Water	K1396-10	pH	Cool 4 deg C	PSEG01	K31	23937	02/06/2019	9040C
02/06/2019	Water	K1396-11	pH	Cool 4 deg C	PSEG01	K31	23976	02/06/2019	9040C
02/06/2019	Water	K1396-12	pH	Cool 4 deg C	PSEG01	K31	23923	02/06/2019	9040C

Date/Time 02-06-19 5:23 PM

Received by: Jo

Relinquished by: ✓

Date/Time 02-06-19 6:35 PM

Received by: cy

Relinquished by: JS