



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB100697

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	02/06/2019	17:20
2	CAL2	1	Water	NA	NA	22.1	7.01	02/06/2019	17:21
3	CAL3	1	Water	NA	NA	22.1	10.03	02/06/2019	17:22
4	ICV	1	Water	NA	NA	22.1	7.00	02/06/2019	17:24
5	CCV1	1	Water	NA	NA	22.2	2.02	02/06/2019	17:25
6	K1389-22	1	Water	NA	NA	10.2	6.60	02/06/2019	17:26
7	K1389-22DUP	1	Water	NA	NA	10.3	6.61	02/06/2019	17:27
8	CCV2	1	Water	NA	NA	22.1	12.01	02/06/2019	17:28

2B100697

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrsoivity k1389

WorkList ID : 122020

Date : 02-06-2019 15:09:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2019	Water	✓ K1389-22	Corrosivity	Cool 4 deg C	PSEG01	K31	M-7	02/06/2019	9040C

Date/Time 02-06-19 5.25 PM
Received by: Jd
Relinquished by: Vm

Date/Time 02-06-19 6.35 PM
Received by:
Relinquished by: Jd